

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060722\
 Data File : PL075587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2022 09:16
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 07:09:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 07 02:04:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.642	4.640	13254292	30922118	19.639	18.426
28)	SA Decachlor...	11.637	10.383	12749099	27321666	20.824	19.392
Target Compounds							
7)	B delta-BHC	0.000	6.406	0	402455	N.D.	0.188 #
10)	B gamma-Chl...	8.382	0.000	825993	0	1.064	N.D. #
11)	B alpha-Chl...	0.000	7.390f	0	220781	N.D.	0.121 #
15)	B Endosulfa...	9.302	0.000	294019	0	0.472	N.D. #
21)	B Endrin ke...	10.185f	0.000	775535	0	1.103	N.D. #
25)	Chlordane-3	8.382	0.000	825993	0	8.000	N.D. #
26)	Chlordane-4	0.000	7.390f	0	220781	N.D.	1.199 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060722\
Data File : PL075587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2022 09:16
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 07:09:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
Quant Title : GC Extractables
QLast Update : Tue Jun 07 02:04:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 5.642 min
Delta R.T.: 0.001 min
Response: 13254292
Conc: 19.64 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.640 min
Delta R.T.: 0.002 min
Response: 30922118
Conc: 18.43 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 7.145 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 6.406 min
Delta R.T.: 0.006 min
Response: 402455
Conc: 0.19 ng/ml



#10 gamma-Chlordane


R.T.: 8.382 min
Delta R.T.: -0.008 min
Response: 825993
Conc: 1.06 ng/ml

Instrument :
ECD_L
ClientSampleId :




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T. : 7.344 min
Response: 0
Conc: N.D.




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 8.472 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 7.390 min
Delta R.T.: -0.022 min
Response: 220781
Conc: 0.12 ng/ml



#15 Endosulfan II


R.T.: 9.302 min
Delta R.T.: 0.009 min
Response: 294019
Conc: 0.47 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.354 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 10.185 min
Delta R.T.: 0.016 min
Response: 775535
Conc: 1.10 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 9.290 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 8.382 min
Delta R.T.: -0.007 min
Response: 825993
Conc: 8.00 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 7.343 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 8.472 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 7.390 min
Delta R.T.: -0.021 min
Response: 220781
Conc: 1.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.637 min
Delta R.T.: 0.001 min
Response: 12749099
Conc: 20.82 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 10.383 min
Delta R.T.: 0.000 min
Response: 27321666
Conc: 19.39 ng/ml