

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060922\
 Data File : PL075677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 18:06
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 23:51:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 07 02:04:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.641	4.636	35307497	79096841	52.316	47.132
28)	SA Decachlor...	11.638	10.383	30934679	59098363	50.528	41.945
Target Compounds							
2)	A alpha-BHC	6.225	5.339	50803540	115.3E6	54.049	50.912
3)	MA gamma-BHC...	6.624	5.760	47322307	107.0E6	53.914	49.361
4)	MA Heptachlor	7.279	6.168	46993058	101.3E6	51.105	47.501
5)	MB Aldrin	7.653	6.492	44634015	96650965	54.362	49.212
6)	B beta-BHC	6.876	6.138	21443340	47208682	52.248	48.096
7)	B delta-BHC	7.147	6.399	44687548	105.0E6	58.075	49.069
8)	B Heptachlo...	8.121	7.066	41590204	88610888	53.804	48.425
9)	A Endosulfan I	8.530	7.468	37281636	80597214	53.283	48.188
10)	B gamma-Chl...	8.392	7.343	40941019	89069863	52.721	48.579
11)	B alpha-Chl...	8.473	7.412	40258546	87696713	52.791	48.030
12)	B 4,4'-DDE	8.659	7.624	34907171	78265102	52.744	48.953
13)	MA Dieldrin	8.820	7.753	37457747	82562783	51.849	48.779
14)	MA Endrin	9.062	8.044	30268945	70505079	51.051	51.572
15)	B Endosulfa...	9.295	8.354	30490579	65940221	48.998	46.013
16)	A 4,4'-DDD	9.201	8.208	29186659	63539466	53.343	48.321
17)	MA 4,4'-DDT	9.521	8.469	29339864	62416322	45.802	44.400
18)	B Endrin al...	9.430	8.542	24053543	50062500	46.964	44.447
19)	B Endosulfa...	9.669	8.772	31528578	67554922	50.333	45.994
20)	A Methoxychlor	10.006	9.061	15854807	34849380	44.291	42.779
21)	B Endrin ke...	10.170	9.290	35022275	72495847	49.801	44.910
22)	Mirex	10.644	9.477	26940790	57614043	47.870	44.866

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060922\
Data File : PL075677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 18:06
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 23:51:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
Quant Title : GC Extractables
QLast Update : Tue Jun 07 02:04:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

