

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060923\
 Data File : PL083357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2023 13:44
 Operator : AR\AJ
 Sample : 03069--02 (Sig #1); 03069-02 (Sig #2)
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 O3069--02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 21:28:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	2.632	63361357	55897829	21.559	21.532
28)	SA Decachlor...	8.729	7.734	44108547	52827148	21.031	21.245

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060923\
Data File : PL083357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2023 13:44
Operator : AR\AJ
Sample : 03069--02 (Sig #1); 03069-02 (Sig #2)
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
O3069--02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 21:28:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.313 min
Delta R.T.: -0.010 min
Response: 63361357
Conc: 21.56 ng/ml

Instrument :
ECD_L
ClientSampleId :
O3069--02




#1 Tetrachloro-m-xylene

R.T.: 2.632 min
Delta R.T.: -0.007 min
Response: 55897829
Conc: 21.53 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.729 min
Delta R.T.: -0.009 min
Response: 44108547
Conc: 21.03 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.734 min
Delta R.T.: -0.009 min
Response: 52827148
Conc: 21.25 ng/ml