

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061020\
 Data File : PL059224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2020 12:24
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 14:32:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060320.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 04 08:16:22 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.520	4.072	90090672	128.2E6	52.171	49.263
28)	SA Decachlor...	8.292	9.180	79773758	109.8E6	48.513	46.301
Target Compounds							
2)	A alpha-BHC	3.961	4.465	133.1E6	186.8E6	53.174	48.600
3)	MA gamma-BHC...	4.246	4.752	115.1E6	175.2E6	52.498	48.008
4)	MA Heptachlor	4.552	5.265	104.4E6	161.0E6	50.168	46.698
5)	MB Aldrin	4.802	5.573	91853549	149.2E6	51.573	47.409
6)	B beta-BHC	4.495	4.919	47382273	73740901	49.865	48.763
7)	B delta-BHC	4.702	5.139	104.0E6	163.7E6	52.237	58.358
8)	B Heptachlo...	5.245	5.961	85890565	139.9E6	50.283	50.499
9)	A Endosulfan I	5.586	6.319	78180359	128.9E6	49.980	46.069
10)	B gamma-Chl...	5.473	6.196	84286834	140.1E6	50.336	46.839
11)	B alpha-Chl...	5.532	6.269	83802875	140.1E6	50.234	46.109
12)	B 4,4'-DDE	5.695	6.423	76291316	128.5E6	50.169	46.171
13)	MA Dieldrin	5.830	6.578	80909012	133.7E6	49.643	45.789
14)	MA Endrin	6.087	6.796	67496831	103.2E6	44.548	41.426
15)	B Endosulfa...	6.359	7.005	75795925	118.0E6	49.040	45.302
16)	A 4,4'-DDD	6.211	6.914	68220361	110.1E6	51.605	47.438
17)	MA 4,4'-DDT	6.453	7.214	60262139	96608094	47.773	42.424
18)	B Endrin al...	6.528	7.129	64583140	103.0E6	49.049	45.060
19)	B Endosulfa...	6.741	7.352	73391768	112.5E6	49.265	44.913
20)	A Methoxychlor	6.999	7.671	35674216	55815892	46.443	42.328
21)	B Endrin ke...	7.233	7.825	88509781	125.7E6	50.974	47.896
22)	Mirex	7.426	8.282	66007403	90981723	47.826	46.120

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061020\
Data File : PL059224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2020 12:24
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 14:32:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060320.M
Quant Title : GC Extractables
QLast Update : Thu Jun 04 08:16:22 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

