

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061119\
 Data File : PL049416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2019 12:42
 Operator : SM\AJ
 Sample : WATER IDOC 02
 Misc : FOR RC
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WATER IDOC 02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 01:10:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.936	225.5E6	64617177	20.465	22.366
28)	SA Decachlor...	7.992	8.972	205.3E6	60561465	16.145	18.134
Target Compounds							
2)	A alpha-BHC	3.753	4.324	931.7E6	216.5E6	51.714	54.018
3)	MA gamma-BHC...	4.029	4.608	875.1E6	203.4E6	51.635	54.320
4)	MA Heptachlor	4.314	5.111	922.5E6	216.1E6	52.095	53.777
5)	MB Aldrin	4.552	5.414	889.7E6	201.7E6	52.110	54.820
6)	B beta-BHC	4.283	4.783	364.9E6	91903767	52.515	53.356
7)	B delta-BHC	4.476	4.997	779.2E6	165.3E6	48.214	53.880
8)	B Heptachlo...	4.989	5.800	795.4E6	183.4E6	53.636	56.092
9)	A Endosulfan I	5.319	6.156	751.0E6	168.5E6	51.317	52.315
10)	B gamma-Chl...	5.211	6.034	826.5E6	186.0E6	51.776	53.761
11)	B alpha-Chl...	5.268	6.106	798.1E6	183.4E6	51.091	52.872
12)	B 4,4'-DDE	5.435	6.266	791.7E6	175.6E6	50.616	53.490
13)	MA Dieldrin	5.558	6.412	794.0E6	177.1E6	51.930	52.794
14)	MA Endrin	5.811	6.629	731.7E6	153.9E6	50.733	51.585
15)	B Endosulfa...	6.086	6.840	658.4E6	147.7E6	49.860	48.706
16)	A 4,4'-DDD	5.946	6.756	601.5E6	132.4E6	50.546	50.347
17)	MA 4,4'-DDT	6.180	7.053	622.7E6	138.8E6	50.576	49.503
18)	B Endrin al...	6.252	6.965	536.6E6	127.6E6	49.031	49.492
19)	B Endosulfa...	6.464	7.187	582.6E6	141.5E6	44.871	46.548
20)	A Methoxychlor	6.725	7.513	288.2E6	75729874	44.934	45.680
21)	B Endrin ke...	6.946	7.656	621.1E6	146.3E6	43.640	46.005
22)	Mirex	7.120	8.105	475.4E6	123.9E6	45.122	47.718

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061119\
Data File : PL049416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2019 12:42
Operator : SM\AJ
Sample : WATER IDOC 02
Misc : FOR RC
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
WATER IDOC 02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 01:10:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

