

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061223\
 Data File : PL083382.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2023 17:40
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/13/2023
 Supervised By : Ankita Jodhani 06/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 22:30:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	2.632	144.4E6	126.7E6	49.145	48.801
28)	SA Decachlor...	8.728	7.733	112.2E6	124.4E6	53.509	50.023
Target Compounds							
2)	A alpha-BHC	3.762	3.127	220.5E6	182.7E6	49.676	48.975
3)	MA gamma-BHC...	4.090	3.452	205.2E6	168.1E6	49.538	50.421m
4)	MA Heptachlor	4.667	3.783	195.1E6	174.1E6	51.046	47.552
5)	MB Aldrin	5.004	4.059	182.4E6	159.0E6	51.840	48.866
6)	B beta-BHC	4.295	3.755	89424557	68658358	47.497	48.010
7)	B delta-BHC	4.535	3.979	198.7E6	164.0E6	57.776	50.069
8)	B Heptachlo...	5.430	4.558	164.7E6	145.8E6	51.501	49.131m
9)	A Endosulfan I	5.811	4.927	152.0E6	136.6E6	50.642	48.873
10)	B gamma-Chl...	5.685	4.809	164.6E6	146.0E6	51.801	48.616
11)	B alpha-Chl...	5.762	4.871	162.5E6	146.0E6	50.621	48.642
12)	B 4,4'-DDE	5.941	5.061	145.7E6	124.5E6	52.868	51.241m
13)	MA Dieldrin	6.085	5.189	154.6E6	138.3E6	51.401	49.004m
14)	MA Endrin	6.312	5.463	133.0E6	120.5E6	49.478	47.504m
15)	B Endosulfa...	6.535	5.762	126.9E6	115.6E6	47.790	46.731
16)	A 4,4'-DDD	6.458	5.615	119.2E6	101.6E6	54.847	52.806m
17)	MA 4,4'-DDT	6.768	5.863	123.9E6	107.5E6	49.511	48.194m
18)	B Endrin al...	6.664	5.941	103.9E6	93879729	51.663m	47.475
19)	B Endosulfa...	6.900	6.164	132.0E6	119.7E6	51.488	48.880
20)	A Methoxychlor	7.250	6.443	65945663	59140851	49.869	46.504
21)	B Endrin ke...	7.379	6.667	135.9E6	133.7E6	53.273	49.460m
22)	Mirex	7.842	6.837	105.3E6	112.9E6	50.246	48.242m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061223\
Data File : PL083382.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2023 17:40
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/13/2023

Supervised By :Ankita Jodhani 06/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 22:30:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

