

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061322\
 Data File : PL075747.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2022 22:02
 Operator : AR\AJ
 Sample : N3304-01 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WC-14A-1A-10

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/14/2022
 Supervised By : Ankita Jodhani 06/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 23:24:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 07 02:04:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.642	4.638	826900	1523697	1.225	0.908 #
28) SA Decachlor...	11.632	10.381	894859	1596154	1.462m	1.133
Target Compounds						
16) A 4,4'-DDD	9.199	8.205	614443	1260448	1.123m	0.959

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061322\
Data File : PL075747.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2022 22:02
Operator : AR\AJ
Sample : N3304-01 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
WC-14A-1A-10


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/14/2022
Supervised By :Ankita Jodhani 06/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 23:24:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
Quant Title : GC Extractables
QLast Update : Tue Jun 07 02:04:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 826900
Conc: 1.23 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-14A-1A-10


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/14/2022
Supervised By :Ankita Jodhani 06/14/2022




#1 Tetrachloro-m-xylene

R.T.: 4.638 min
Delta R.T.: 0.000 min
Response: 1523697
Conc: 0.91 ng/ml




#16 4,4'-DDD

R.T.: 9.199 min
Delta R.T.: 0.000 min
Response: 614443
Conc: 1.12 ng/ml m



#16 4,4'-DDD

R.T.: 8.205 min
Delta R.T.: -0.003 min
Response: 1260448
Conc: 0.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.632 min
Delta R.T.: -0.004 min
Response: 894859
Conc: 1.46 ng/ml

Instrument :

ECD_L

ClientSampleId :


WC-14A-1A-10

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/14/2022

Supervised By :Ankita Jodhani 06/14/2022



#28 Decachlorobiphenyl

R.T.: 10.381 min
Delta R.T.: -0.003 min
Response: 1596154
Conc: 1.13 ng/ml