

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061419\
 Data File : PL049491.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2019 12:10
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 6/17/2019 9:28:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 15:09:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.936	521.4E6	146.6E6	47.316	50.739
28)	SA Decachlor...	7.996	8.973	540.6E6	149.2E6	42.516	44.660
Target Compounds							
2)	A alpha-BHC	3.756	4.325	879.6E6	215.3E6	48.823	53.706
3)	MA gamma-BHC...	4.031	4.608	793.5E6	198.7E6	46.820m	53.083
4)	MA Heptachlor	4.316	5.111	835.8E6	202.6E6	47.199	50.421
5)	MB Aldrin	4.555	5.414	799.7E6	183.5E6	46.837	49.868
6)	B beta-BHC	4.286	4.783	332.6E6	88740283	47.863	51.520
7)	B delta-BHC	4.480	4.997	832.6E6	170.7E6	51.519	55.648
8)	B Heptachlor...	4.992	5.800	734.2E6	158.2E6	49.508	48.405
9)	A Endosulfan I	5.323	6.156	698.8E6	157.8E6	47.752	48.985
10)	B gamma-Chl...	5.214	6.035	773.0E6	172.0E6	48.420	49.727
11)	B alpha-Chl...	5.271	6.107	748.0E6	171.6E6	47.883	49.450
12)	B 4,4'-DDE	5.439	6.267	731.8E6	163.0E6	46.784	49.643
13)	MA Dieldrin	5.562	6.413	750.9E6	166.1E6	49.106	49.524
14)	MA Endrin	5.814	6.630	702.2E6	146.4E6	48.688	49.061
15)	B Endosulfa...	6.090	6.842	652.4E6	147.9E6	49.407	48.758
16)	A 4,4'-DDD	5.949	6.758	578.7E6	125.3E6	48.631	47.643
17)	MA 4,4'-DDT	6.184	7.054	600.5E6	134.8E6	48.771	48.065
18)	B Endrin al...	6.257	6.966	527.9E6	125.2E6	48.236	48.556
19)	B Endosulfa...	6.468	7.188	573.7E6	136.9E6	44.190	45.025
20)	A Methoxychlor	6.729	7.515	287.7E6	75157451	44.860	45.334
21)	B Endrin ke...	6.949	7.657	629.6E6	145.1E6	44.238	45.604
22)	Mirex	7.124	8.107	443.2E6	113.0E6	42.063	43.497

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061419\
Data File : PL049491.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2019 12:10
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
6/17/2019 9:28:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 15:09:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

