

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061423\
 Data File : PL083437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2023 18:20
 Operator : AR\AJ
 Sample : 03157-05MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SU-622-COMP-01MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 22:38:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	2.632	56025194	46460697	19.063	17.896
2)	SA Decachlor...	8.731	7.733	46808798	52794695	22.318	21.232
Target Compounds							
2)	A alpha-BHC	3.762	3.127	183.1E6	154.7E6	41.244	41.459
3)	MA gamma-BHC...	4.090	3.453	175.3E6	142.0E6	42.323	42.590
4)	MA Heptachlor	4.668	3.783	168.8E6	153.6E6	44.163	41.956
5)	MB Aldrin	5.004	4.058	158.7E6	138.7E6	45.106	42.640
6)	B beta-BHC	4.295	3.754	79989756	61262595	42.486	42.838
7)	B delta-BHC	4.535	3.978	125.1E6	100.4E6	36.382	30.645
8)	B Heptachlo...	5.430	4.559	147.7E6	123.6E6	46.179	41.662
9)	A Endosulfan I	5.811	4.926	137.3E6	123.8E6	45.753	44.313
10)	B gamma-Chl...	5.685	4.808	153.2E6	132.8E6	48.233	44.203
11)	B alpha-Chl...	5.763	4.870	147.0E6	132.2E6	45.803	44.046
12)	B 4,4'-DDE	5.942	5.061	130.9E6	110.7E6	47.480	45.589
13)	MA Dieldrin	6.086	5.190	138.5E6	118.1E6	46.062	41.831
14)	MA Endrin	6.313	5.464	119.3E6	101.5E6	44.379	39.992
15)	B Endosulfa...	6.536	5.761	114.4E6	105.8E6	42.649	42.778
16)	A 4,4'-DDD	6.458	5.615	108.9E6	91660273	50.117	47.659
17)	MA 4,4'-DDT	6.769	5.863	109.5E6	91910373	43.735	41.196
18)	B Endrin al...	6.666	5.941	86457142	83353434	43.000	42.152
19)	B Endosulfa...	6.900	6.164	113.3E6	103.8E6	44.198	42.404
20)	A Methoxychlor	7.251	6.443	56478607	52862813	42.710	41.567
21)	B Endrin ke...	7.380	6.668	121.1E6	119.0E6	47.475	44.016
22)	Mirex	7.844	6.838	95352550	97621833	45.492	41.722

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061423\
Data File : PL083437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2023 18:20
Operator : AR\AJ
Sample : 03157-05MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SU-622-COMP-01MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 22:38:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

