

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061621\
 Data File : PL067518.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2021 10:40
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 04:48:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 11 02:00:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	5.529	9712119	17894160	17.665	18.981
28)	SA Decachlor...	10.222	11.428	14120806	21593908	19.925	21.679
Target Compounds							
2)	A alpha-BHC	5.166	0.000	369171	0	0.502	N.D. #
4)	MA Heptachlor	5.988f	0.000	645920	0	0.856	N.D. #
6)	B beta-BHC	5.988	0.000	645920	0	1.805	N.D. #
15)	B Endosulfa...	0.000	9.179f	0	428088	N.D.	0.537 #
24)	Chlordane-2	6.486	0.000	319263	0	11.922	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061621\
Data File : PL067518.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2021 10:40
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 04:48:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
Quant Title : GC Extractables
QLast Update : Fri Jun 11 02:00:12 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: -0.003 min
Response: 9712119
Conc: 17.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 5.529 min
Delta R.T.: -0.002 min
Response: 17894160
Conc: 18.98 ng/ml




#2 alpha-BHC

R.T.: 5.166 min
Delta R.T.: -0.014 min
Response: 369171
Conc: 0.50 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 6.112 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.988 min
Delta R.T.: -0.024 min
Response: 645920
Conc: 0.86 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 7.155 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 5.988 min
Delta R.T.: 0.005 min
Response: 645920
Conc: 1.81 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 6.762 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T. : 8.190 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
I.BLK




#15 Endosulfan II

R.T.: 9.179 min
Delta R.T.: 0.021 min
Response: 428088
Conc: 0.54 ng/ml




#24 Chlordane-2

R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 319263
Conc: 11.92 ng/ml



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T. : 7.503 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.222 min
Delta R.T.: -0.003 min
Response: 14120806
Conc: 19.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 11.428 min
Delta R.T.: 0.000 min
Response: 21593908
Conc: 21.68 ng/ml