

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL061721\  
 Data File : PL067608.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 18 Jun 2021 04:12  
 Operator : AR\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_L  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 18 04:45:07 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL061021.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Jun 11 02:00:12 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

|                             | Compound        | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|-----------------|--------|--------|----------|----------|--------|--------|
| -----                       |                 |        |        |          |          |        |        |
| System Monitoring Compounds |                 |        |        |          |          |        |        |
| 1)                          | SA Tetrachlo... | 4.478  | 5.528  | 11395665 | 18571472 | 20.727 | 19.700 |
| 28)                         | SA Decachlor... | 10.218 | 11.425 | 10897577 | 17671611 | 15.377 | 17.742 |
| Target Compounds            |                 |        |        |          |          |        |        |
| 2)                          | A alpha-BHC     | 5.181  | 0.000  | 342528   | 0        | 0.466  | N.D. # |
| 4)                          | MA Heptachlor   | 5.987f | 0.000  | 558524   | 0        | 0.740  | N.D. # |
| 6)                          | B beta-BHC      | 5.987  | 0.000  | 558524   | 0        | 1.561  | N.D. # |
| 14)                         | MA Endrin       | 7.892  | 0.000  | 682615   | 0        | 1.041  | N.D. # |
| 24)                         | Chlordane-2     | 6.485  | 0.000  | 261305   | 0        | 9.758  | N.D. # |
| -----                       |                 |        |        |          |          |        |        |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_L\Data\PL061721\  
Data File : PL067608.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 18 Jun 2021 04:12  
Operator : AR\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_L  
ClientSampled :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 18 04:45:07 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_L\methods\PL061021.M  
Quant Title : GC Extractables  
QLast Update : Fri Jun 11 02:00:12 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min  
Delta R.T.: -0.003 min  
Response: 11395665  
Conc: 20.73 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 5.528 min  
Delta R.T.: -0.003 min  
Response: 18571472  
Conc: 19.70 ng/ml




#2 alpha-BHC

R.T.: 5.181 min  
Delta R.T.: 0.000 min  
Response: 342528  
Conc: 0.47 ng/ml



#2 alpha-BHC


R.T.: 0.000 min  
Exp R.T. : 6.112 min  
Response: 0  
Conc: N.D.



#4 Heptachlor


R.T.: 5.987 min  
Delta R.T.: -0.025 min  
Response: 558524  
Conc: 0.74 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK




#4 Heptachlor

R.T.: 0.000 min  
Exp R.T. : 7.155 min  
Response: 0  
Conc: N.D.




#6 beta-BHC

R.T.: 5.987 min  
Delta R.T.: 0.004 min  
Response: 558524  
Conc: 1.56 ng/ml



#6 beta-BHC


R.T.: 0.000 min  
Exp R.T. : 6.762 min  
Response: 0  
Conc: N.D.



#14 Endrin


R.T.: 7.892 min  
Delta R.T.: 0.014 min  
Response: 682615  
Conc: 1.04 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK




#14 Endrin

R.T.: 0.000 min  
Exp R.T. : 8.924 min  
Response: 0  
Conc: N.D.




#24 Chlordane-2

R.T.: 6.485 min  
Delta R.T.: -0.001 min  
Response: 261305  
Conc: 9.76 ng/ml



#24 Chlordane-2


R.T.: 0.000 min  
Exp R.T. : 7.503 min  
Response: 0  
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.218 min  
Delta R.T.: -0.006 min  
Response: 10897577  
Conc: 15.38 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :  
I.BLK



#28 Decachlorobiphenyl

R.T.: 11.425 min  
Delta R.T.: -0.003 min  
Response: 17671611  
Conc: 17.74 ng/ml