

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061819\
 Data File : PL049584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2019 11:54
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 01:03:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.936	527.1E6	142.0E6	47.836	49.154
28)	SA Decachlor...	7.994	8.973	542.4E6	145.2E6	42.656	43.476
Target Compounds							
2)	A alpha-BHC	3.756	4.325	890.7E6	209.3E6	49.438	52.211
3)	MA gamma-BHC...	4.032	4.608	850.2E6	194.2E6	50.164	51.879
4)	MA Heptachlor	4.317	5.111	861.8E6	198.8E6	48.664	49.483
5)	MB Aldrin	4.555	5.414	827.8E6	179.1E6	48.480	48.684
6)	B beta-BHC	4.286	4.784	342.9E6	89519358	49.342	51.972
7)	B delta-BHC	4.480	4.998	856.9E6	175.4E6	53.025	57.178
8)	B Heptachlor...	4.992	5.801	746.9E6	161.7E6	50.364	49.467
9)	A Endosulfan I	5.322	6.156	734.6E6	153.9E6	50.198	47.797
10)	B gamma-Chl...	5.214	6.035	808.5E6	169.0E6	50.646	48.843
11)	B alpha-Chl...	5.270	6.106	784.8E6	167.1E6	50.234	48.178
12)	B 4,4'-DDE	5.438	6.267	777.8E6	159.2E6	49.728	48.505
13)	MA Dieldrin	5.561	6.413	786.9E6	162.2E6	51.461	48.361
14)	MA Endrin	5.814	6.629	747.6E6	144.2E6	51.835	48.311
15)	B Endosulfa...	6.089	6.842	678.6E6	143.2E6	51.389	47.212
16)	A 4,4'-DDD	5.949	6.757	615.9E6	123.7E6	51.755	47.023
17)	MA 4,4'-DDT	6.183	7.054	634.3E6	131.1E6	51.515	46.768
18)	B Endrin al...	6.256	6.965	552.8E6	122.1E6	50.507	47.351
19)	B Endosulfa...	6.467	7.188	600.6E6	133.9E6	46.259	44.023
20)	A Methoxychlor	6.728	7.515	302.3E6	72857347	47.129	43.947
21)	B Endrin ke...	6.948	7.657	658.6E6	141.5E6	46.280	44.497
22)	Mirex	7.122	8.106	449.7E6	110.1E6	42.677	42.376

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL061819\
Data File : PL049584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2019 11:54
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 01:03:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

