

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062022\
 Data File : PL075873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 19:27
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/05/2022
 Supervised By : Ankita Jodhani 07/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 02:45:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 07 02:04:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.642	4.636	37613039	92675364	55.732	55.223
28)	SA Decachlor...	11.635	10.377	27026900	57194792	44.146	40.594
Target Compounds							
2)	A alpha-BHC	6.226	5.338	54135593	131.3E6	57.593	57.997
3)	MA gamma-BHC...	6.624	5.758	50145745	121.2E6	57.131	55.882
4)	MA Heptachlor	7.278	6.166	48870325	116.9E6	53.146	54.843
5)	MB Aldrin	7.652	6.489	45271730	108.2E6	55.139	55.068
6)	B beta-BHC	6.876	6.137	22457980	51445861	54.720	52.413
7)	B delta-BHC	7.146	6.397	41419510	118.8E6	53.828	55.494
8)	B Heptachlo...	8.120	7.063	41881816	98195517	54.182	53.663
9)	A Endosulfan I	8.528	7.464	36412473	89550495	52.041	53.541
10)	B gamma-Chl...	8.389	7.339	38563839	98956352	49.660m	53.971
11)	B alpha-Chl...	8.471	7.408	39731476	97567154	52.100	53.436
12)	B 4,4'-DDE	8.658	7.620	33850655	87258079	51.147	54.578
13)	MA Dieldrin	8.818	7.749	36097659	91243984	49.967	53.908
14)	MA Endrin	9.060	8.039	29654873	69855551	50.016	51.097
15)	B Endosulfa...	9.293	8.349	29218298	71256199	46.953	49.723
16)	A 4,4'-DDD	9.199	8.203	27219920	67683501	49.748	51.472
17)	MA 4,4'-DDT	9.519	8.463	28977456	69283552	45.236	49.285
18)	B Endrin al...	9.428	8.537	23483954	53325632	45.852	47.344
19)	B Endosulfa...	9.667	8.767	29541377	68284763	47.161	46.491
20)	A Methoxychlor	10.004	9.056	15542505	36872611	43.418	45.262
21)	B Endrin ke...	10.168	9.285	31559824	72014099	44.878	44.611
22)	Mirex	10.642	9.471	24256508	54618851	43.100	42.534

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062022\
 Data File : PL075873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 19:27
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 07/05/2022

Supervised By :Ankita Jodhani 07/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 02:45:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 07 02:04:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

