

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062023\
 Data File : PL083608.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2023 22:00
 Operator : AR\AJ
 Sample : 03328-13
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 COMPOSIT-SW02-0-8

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/21/2023
 Supervised By :Ankita Jodhani 06/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 07:18:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061723.M
 Quant Title : GC Extractables
 QLast Update : Sat Jun 17 12:34:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.451	2.752	25733894	28651048	11.643m	12.300
28) SA Decachlor...	8.989	7.959	31872702	36932480	16.726	15.037
Target Compounds						
16) A 4,4'-DDD	6.648	5.806	10540288	1928336	6.173m	0.874m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062023\
Data File : PL083608.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2023 22:00
Operator : AR\AJ
Sample : 03328-13
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
COMPOSIT-SW02-0-8


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/21/2023
Supervised By : Ankita Jodhani 06/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 07:18:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061723.M
Quant Title : GC Extractables
QLast Update : Sat Jun 17 12:34:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.451 min
Delta R.T.: 0.000 min
Response: 25733894
Conc: 11.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
COMPOSIT-SW02-0-8


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/21/2023
Supervised By :Ankita Jodhani 06/21/2023




#1 Tetrachloro-m-xylene

R.T.: 2.752 min
Delta R.T.: 0.000 min
Response: 28651048
Conc: 12.30 ng/ml




#16 4,4'-DDD

R.T.: 6.648 min
Delta R.T.: 0.009 min
Response: 10540288
Conc: 6.17 ng/ml m



#16 4,4'-DDD

R.T.: 5.806 min
Delta R.T.: -0.005 min
Response: 1928336
Conc: 0.87 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.989 min
Delta R.T.: -0.002 min
Response: 31872702
Conc: 16.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
COMPOSIT-SW02-0-8

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/21/2023
Supervised By :Ankita Jodhani 06/21/2023



#28 Decachlorobiphenyl

R.T.: 7.959 min
Delta R.T.: -0.005 min
Response: 36932480
Conc: 15.04 ng/ml