

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062121\
 Data File : PL067655.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2021 13:28
 Operator : AR\AJ
 Sample : M2709-15
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SPLP-AMND-COMP-M

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 16:31:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 11 02:00:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	5.529	12213623	21720425	22.215	23.040
28)	SA Decachlor...	10.220	11.427	14935478	22080130	21.075	22.168
Target Compounds							
2)	A alpha-BHC	0.000	6.088f	0	521075	N.D.	0.421 #
5)	MB Aldrin	6.337	0.000	1044883	0	1.420	N.D. #
15)	B Endosulfa...	0.000	9.178f	0	296142	N.D.	0.371 #
27)	Chlordane-5	0.000	9.231	0	286624	N.D.	9.857 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062121\
Data File : PL067655.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2021 13:28
Operator : AR\AJ
Sample : M2709-15
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SPLP-AMND-COMP-M

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 16:31:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
Quant Title : GC Extractables
QLast Update : Fri Jun 11 02:00:12 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: -0.003 min
Response: 12213623
Conc: 22.22 ng/ml

Instrument :
ECD_L
ClientSampleId :
SPLP-AMND-COMP-M




#1 Tetrachloro-m-xylene

R.T.: 5.529 min
Delta R.T.: -0.002 min
Response: 21720425
Conc: 23.04 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 5.181 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 6.088 min
Delta R.T.: -0.024 min
Response: 521075
Conc: 0.42 ng/ml



#5 Aldrin


R.T.: 6.337 min
Delta R.T.: 0.003 min
Response: 1044883
Conc: 1.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
SPLP-AMND-COMP-M




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 7.525 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.190 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 9.178 min
Delta R.T.: 0.019 min
Response: 296142
Conc: 0.37 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 8.191 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
SPLP-AMND-COMP-M




#27 Chlordane-5

R.T.: 9.231 min
Delta R.T.: 0.005 min
Response: 286624
Conc: 9.86 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.004 min
Response: 14935478
Conc: 21.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.427 min
Delta R.T.: -0.001 min
Response: 22080130
Conc: 22.17 ng/ml