

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062322\
 Data File : PL075963.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2022 03:17
 Operator : AR\AJ
 Sample : N3431-13
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WC-14B-2-4

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/24/2022
 Supervised By : Ankita Jodhani 06/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 05:21:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.642	4.637	6167739	14228189	8.155	7.502
28) SA Decachlor...	11.634	10.376	7005191	14224843	11.132m	9.262m
Target Compounds						
12) B 4,4'-DDE	8.660	7.619	3338201	4590604	4.939	2.705 #
16) A 4,4'-DDD	9.200	8.203	2622748	5018582	4.625	3.618

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062322\
Data File : PL075963.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2022 03:17
Operator : AR\AJ
Sample : N3431-13
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
WC-14B-2-4


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/24/2022
Supervised By : Ankita Jodhani 06/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 05:21:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.642 min
Delta R.T.: 0.002 min
Response: 6167739
Conc: 8.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-14B-2-4


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/24/2022
Supervised By :Ankita Jodhani 06/24/2022




#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: 0.003 min
Response: 14228189
Conc: 7.50 ng/ml




#12 4,4'-DDE

R.T.: 8.660 min
Delta R.T.: 0.004 min
Response: 3338201
Conc: 4.94 ng/ml



#12 4,4'-DDE

R.T.: 7.619 min
Delta R.T.: 0.001 min
Response: 4590604
Conc: 2.70 ng/ml




#16 4,4'-DDD

R.T.: 9.200 min
Delta R.T.: 0.002 min
Response: 2622748
Conc: 4.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-14B-2-4


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/24/2022
Supervised By :Ankita Jodhani 06/24/2022




#16 4,4'-DDD

R.T.: 8.203 min
Delta R.T.: 0.000 min
Response: 5018582
Conc: 3.62 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.634 min
Delta R.T.: 0.000 min
Response: 7005191
Conc: 11.13 ng/ml m



#28 Decachlorobiphenyl

R.T.: 10.376 min
Delta R.T.: 0.000 min
Response: 14224843
Conc: 9.26 ng/ml m