

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062322\
 Data File : PL075964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2022 03:34
 Operator : AR\AJ
 Sample : N3431-19
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WC-14B-2-5

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/24/2022
 Supervised By : Ankita Jodhani 06/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 05:21:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.642	4.637	6065446	14340124	8.020	7.561
28)	SA Decachlor...	11.636	10.378	6885243	14043068	10.942	9.143
Target Compounds							
12)	B 4,4'-DDE	8.657	7.620	2388432	4028476	3.534m	2.373 #
16)	A 4,4'-DDD	9.199	8.202	1800317	2781059	3.175m	2.005m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062322\
Data File : PL075964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2022 03:34
Operator : AR\AJ
Sample : N3431-19
Misc :
ALS Vial : 36 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
WC-14B-2-5


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/24/2022
Supervised By : Ankita Jodhani 06/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 05:21:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.642 min
Delta R.T.: 0.002 min
Response: 6065446
Conc: 8.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-14B-2-5


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/24/2022
Supervised By :Ankita Jodhani 06/24/2022




#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: 0.003 min
Response: 14340124
Conc: 7.56 ng/ml




#12 4,4'-DDE

R.T.: 8.657 min
Delta R.T.: 0.001 min
Response: 2388432
Conc: 3.53 ng/ml m



#12 4,4'-DDE

R.T.: 7.620 min
Delta R.T.: 0.002 min
Response: 4028476
Conc: 2.37 ng/ml




#16 4,4'-DDD

R.T.: 9.199 min
Delta R.T.: 0.000 min
Response: 1800317
Conc: 3.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-14B-2-5


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/24/2022
Supervised By :Ankita Jodhani 06/24/2022




#16 4,4'-DDD

R.T.: 8.202 min
Delta R.T.: 0.000 min
Response: 2781059
Conc: 2.01 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.636 min
Delta R.T.: 0.002 min
Response: 6885243
Conc: 10.94 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.378 min
Delta R.T.: 0.002 min
Response: 14043068
Conc: 9.14 ng/ml