

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062520\
 Data File : PL059506.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2020 15:32
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 17:00:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062220.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 22 14:09:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.519	4.071	88641143	128.6E6	49.324	49.831
2)	SA Decachlor...	8.289	9.182	76834108	104.7E6	45.122	45.744
Target Compounds							
2)	A alpha-BHC	3.960	4.464	131.1E6	189.5E6	50.779	55.618
3)	MA gamma-BHC...	4.245	4.751	115.4E6	177.2E6	50.593	56.611
4)	MA Heptachlor	4.550	5.265	106.4E6	167.1E6	49.902	51.600
5)	MB Aldrin	4.800	5.574	92363588	152.2E6	51.905	49.412
6)	B beta-BHC	4.494	4.920	47128119	74105114	49.488	54.487
7)	B delta-BHC	4.701	5.140	103.7E6	161.8E6	50.820	58.145
8)	B Heptachlo...	5.244	5.961	87302962	147.7E6	58.078	58.834
9)	A Endosulfan I	5.585	6.321	79263273	131.1E6	53.878	48.318
10)	B gamma-Chl...	5.472	6.196	84961145	143.3E6	55.395	48.927
11)	B alpha-Chl...	5.530	6.270	84792992	143.3E6	54.563	48.653
12)	B 4,4'-DDE	5.694	6.425	79919944	130.2E6	55.001	48.227
13)	MA Dieldrin	5.828	6.579	82521261	134.4E6	52.309	47.756
14)	MA Endrin	6.086	6.798	74512032	107.5E6	49.040	43.532
15)	B Endosulfa...	6.358	7.006	76631247	117.9E6	49.623	46.474
16)	A 4,4'-DDD	6.209	6.916	67077898	107.0E6	51.402	48.100
17)	MA 4,4'-DDT	6.452	7.216	64877945	102.1E6	49.204	44.919
18)	B Endrin al...	6.526	7.130	65044463	101.4E6	49.214	46.244
19)	B Endosulfa...	6.739	7.353	73832477	111.5E6	48.629	46.071
20)	A Methoxychlor	6.997	7.673	38349654	58265830	47.320	43.893
21)	B Endrin ke...	7.231	7.826	85100363	119.6E6	48.809	46.308
22)	Mirex	7.424	8.284	66103017	89108476	46.812	45.115

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062520\
Data File : PL059506.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2020 15:32
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 17:00:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062220.M
Quant Title : GC Extractables
QLast Update : Mon Jun 22 14:09:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

