

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062619\
 Data File : PL049870.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2019 14:27
 Operator : SM\AJ
 Sample : PB120927BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB120927BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 00:57:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.323	3.933	229.3E6	65166271	20.334	21.284
28)	SA Decachlor...	7.987	8.967	249.0E6	68597173	19.594	20.434
Target Compounds							
2)	A alpha-BHC	3.750	4.322	943.7E6	223.6E6	50.232	49.957
3)	MA gamma-BHC...	4.026	4.605	891.7E6	211.9E6	50.292	51.080
4)	MA Heptachlor	4.310	5.108	952.5E6	223.4E6	52.404	51.620
5)	MB Aldrin	4.548	5.410	921.6E6	209.7E6	53.370	53.539
6)	B beta-BHC	4.280	4.780	373.4E6	96583540	51.047	49.861
7)	B delta-BHC	4.474	4.994	803.6E6	175.6E6	44.932	44.590
8)	B Heptachlo...	4.985	5.796	842.8E6	189.5E6	52.356	52.108
9)	A Endosulfan I	5.316	6.152	804.4E6	175.2E6	51.987	51.555
10)	B gamma-Chl...	5.207	6.030	879.0E6	191.7E6	51.436	51.432
11)	B alpha-Chl...	5.263	6.102	850.6E6	188.8E6	51.243	51.039
12)	B 4,4'-DDE	5.431	6.262	854.4E6	183.0E6	51.287	52.739
13)	MA Dieldrin	5.555	6.408	859.6E6	181.6E6	51.601	51.189
14)	MA Endrin	5.807	6.625	753.7E6	149.3E6	47.064	47.297
15)	B Endosulfa...	6.082	6.837	729.3E6	156.7E6	50.053	48.636
16)	A 4,4'-DDD	5.942	6.753	689.4E6	136.1E6	54.607	52.060
17)	MA 4,4'-DDT	6.176	7.050	639.7E6	137.3E6	48.183	46.865
18)	B Endrin al...	6.249	6.961	591.0E6	134.9E6	50.034	49.108
19)	B Endosulfa...	6.460	7.184	646.0E6	147.5E6	49.824	49.026
20)	A Methoxychlor	6.721	7.511	300.5E6	75217374	46.108	45.127
21)	B Endrin ke...	6.942	7.652	727.9E6	162.8E6	51.791	52.252
22)	Mirex	7.115	8.101	541.4E6	133.3E6	52.373	52.734

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062619\
Data File : PL049870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2019 14:27
Operator : SM\AJ
Sample : PB120927BSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB120927BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 00:57:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

