

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062620\
 Data File : PL059538.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2020 15:44
 Operator : AJ\MA
 Sample : L3126-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-7-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 23:47:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062220.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 22 14:09:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.519	4.071	28660465	41754755	15.948	16.178
28)	SA Decachlor...	8.290	9.184	21821857	31371144	12.815	13.702
Target Compounds							
2)	A alpha-BHC	3.951	0.000	900398	0	0.349	N.D. #
4)	MA Heptachlor	0.000	5.260	0	894051	N.D.	0.276 #
5)	MB Aldrin	0.000	5.565	0	415836	N.D.	0.135 #
9)	A Endosulfan I	0.000	6.305	0	173057	N.D.	0.064 #
13)	MA Dieldrin	0.000	6.565	0	118873	N.D.	0.042 #
15)	B Endosulfa...	0.000	7.007	0	877278	N.D.	0.346 #
16)	A 4,4'-DDD	0.000	6.923	0	559438	N.D.	0.251 #
17)	MA 4,4'-DDT	6.448	7.188f	237749	294979	0.180	0.130 #
19)	B Endosulfa...	6.736	0.000	168469	0	0.111	N.D. #
20)	A Methoxychlor	6.982	7.687f	352960	-74292	0.436	N.D. #
21)	B Endrin ke...	7.257f	0.000	117320	0	0.067	N.D. #
22)	Mirex	0.000	8.306f	0	467794	N.D.	0.237 #
23)	Chlordane-1	4.422f	0.000	904435	0	15.894	N.D. #
24)	Chlordane-2	0.000	5.565f	0	415836	N.D.	3.788 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062620\
Data File : PL059538.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2020 15:44
Operator : AJ\MA
Sample : L3126-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
TP-7-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 23:47:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062220.M
Quant Title : GC Extractables
QLast Update : Mon Jun 22 14:09:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.519 min
Delta R.T.: 0.002 min
Response: 28660465
Conc: 15.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#1 Tetrachloro-m-xylene

R.T.: 4.071 min
Delta R.T.: 0.000 min
Response: 41754755
Conc: 16.18 ng/ml




#2 alpha-BHC

R.T.: 3.951 min
Delta R.T.: -0.008 min
Response: 900398
Conc: 0.35 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 4.463 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 0.000 min
Exp R.T. : 4.549 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#4 Heptachlor

R.T.: 5.260 min
Delta R.T.: -0.005 min
Response: 894051
Conc: 0.28 ng/ml




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 4.799 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 5.565 min
Delta R.T.: -0.007 min
Response: 415836
Conc: 0.14 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 5.583 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#9 Endosulfan I

R.T.: 6.305 min
Delta R.T.: -0.014 min
Response: 173057
Conc: 0.06 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T.: 5.827 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 6.565 min
Delta R.T.: -0.012 min
Response: 118873
Conc: 0.04 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T. : 6.357 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#15 Endosulfan II

R.T.: 7.007 min
Delta R.T.: 0.002 min
Response: 877278
Conc: 0.35 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 6.209 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.923 min
Delta R.T.: 0.008 min
Response: 559438
Conc: 0.25 ng/ml



#17 4,4'-DDT


R.T.: 6.448 min
Delta R.T.: -0.003 min
Response: 237749
Conc: 0.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#17 4,4'-DDT

R.T.: 7.188 min
Delta R.T.: -0.026 min
Response: 294979
Conc: 0.13 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.736 min
Delta R.T.: -0.003 min
Response: 168469
Conc: 0.11 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T.: 7.352 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 6.982 min
Delta R.T.: -0.014 min
Response: 352960
Conc: 0.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#20 Methoxychlor

R.T.: 7.687 min
Delta R.T.: 0.016 min
Response: -74292
Conc: N.D.




#21 Endrin ketone

R.T.: 7.257 min
Delta R.T.: 0.027 min
Response: 117320
Conc: 0.07 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T.: 7.826 min
Response: 0
Conc: N.D.



#22 Mirex


R.T.: 0.000 min
Exp R.T.: 7.423 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#22 Mirex

R.T.: 8.306 min
Delta R.T.: 0.023 min
Response: 467794
Conc: 0.24 ng/ml




#23 Chlordane-1

R.T.: 4.422 min
Delta R.T.: 0.026 min
Response: 904435
Conc: 15.89 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 5.075 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 4.900 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TP-7-S




#24 Chlordane-2

R.T.: 5.565 min
Delta R.T.: 0.021 min
Response: 415836
Conc: 3.79 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.290 min
Delta R.T.: 0.000 min
Response: 21821857
Conc: 12.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.184 min
Delta R.T.: 0.002 min
Response: 31371144
Conc: 13.70 ng/ml