

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062719\
 Data File : PL049949.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2019 18:50
 Operator : SM\AJ
 Sample : K3157-11
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CL-02-062619-A

Manual Integrations
 APPROVED

Ankita
 6/28/2019 10:16:06 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 01:29:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.322	3.932	121.9E6	30378974	10.810	9.922
28) SA Decachlor...	7.985	8.965	99251468	34480574	7.811	10.271 #
Target Compounds						
10) B gamma-Chl...	5.204	6.027	7562248	3674512	0.443m	0.986m#
11) B alpha-Chl...	5.260	6.101	13718612	7542561	0.826m	2.039m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062719\
Data File : PL049949.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2019 18:50
Operator : SM\AJ
Sample : K3157-11
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
CL-02-062619-A


Manual Integrations
APPROVED

Ankita
6/28/2019 10:16:06 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 01:29:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.322 min
Delta R.T.: -0.002 min
Response: 121907076
Conc: 10.81 ng/ml

Instrument :
ECD_L
ClientSampleId :
CL-02-062619-A


Manual Integrations
APPROVED

Ankita
6/28/2019 10:16:06 AM




#1 Tetrachloro-m-xylene

R.T.: 3.932 min
Delta R.T.: -0.003 min
Response: 30378974
Conc: 9.92 ng/ml




#10 gamma-Chlordane

R.T.: 5.204 min
Delta R.T.: -0.006 min
Response: 7562248
Conc: 0.44 ng/ml m



#10 gamma-Chlordane

R.T.: 6.027 min
Delta R.T.: -0.006 min
Response: 3674512
Conc: 0.99 ng/ml m




#11 alpha-Chlordane

R.T.: 5.260 min
Delta R.T.: -0.007 min
Response: 13718612
Conc: 0.83 ng/ml m

Instrument :
ECD_L
ClientSampleId :
CL-02-062619-A


Manual Integrations
APPROVED

Ankita
6/28/2019 10:16:06 AM




#11 alpha-Chlordane

R.T.: 6.101 min
Delta R.T.: -0.004 min
Response: 7542561
Conc: 2.04 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.985 min
Delta R.T.: -0.007 min
Response: 99251468
Conc: 7.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.965 min
Delta R.T.: -0.008 min
Response: 34480574
Conc: 10.27 ng/ml