

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062723\
 Data File : PL083749.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2023 17:45
 Operator : AR\AJ
 Sample : 03413-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-02-062623

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/28/2023
 Supervised By :Ankita Jodhani 06/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 00:22:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062523.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 26 10:45:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.430	2.693	27717490	33855528	11.156	11.660
28)	SA Decachlor...	8.973	7.881	20168615	26901263	11.809	10.880
Target Compounds							
5)	MB Aldrin	5.155	4.152	183.1E6	205.0E6	53.243	57.308
10)	B gamma-Chl...	5.842	4.912	3444974	2299554	1.023m	0.675m#
11)	B alpha-Chl...	5.926	4.974	5459708	3612230	1.657m	1.067m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062723\
Data File : PL083749.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2023 17:45
Operator : AR\AJ
Sample : 03413-01
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
OR-02-062623


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/28/2023
Supervised By : Ankita Jodhani 06/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 00:22:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062523.M
Quant Title : GC Extractables
QLast Update : Mon Jun 26 10:45:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.430 min
Delta R.T.: 0.000 min
Response: 27717490
Conc: 11.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-02-062623


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/28/2023
Supervised By :Ankita Jodhani 06/28/2023




#1 Tetrachloro-m-xylene

R.T.: 2.693 min
Delta R.T.: 0.000 min
Response: 33855528
Conc: 11.66 ng/ml




#5 Aldrin

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 183104571
Conc: 53.24 ng/ml



#5 Aldrin

R.T.: 4.152 min
Delta R.T.: 0.000 min
Response: 204987559
Conc: 57.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.842 min
Delta R.T.: 0.002 min
Response: 3444974
Conc: 1.02 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-02-062623


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/28/2023
Supervised By :Ankita Jodhani 06/28/2023




#10 gamma-Chlordane

R.T.: 4.912 min
Delta R.T.: 0.000 min
Response: 2299554
Conc: 0.68 ng/ml m




#11 alpha-Chlordane

R.T.: 5.926 min
Delta R.T.: 0.004 min
Response: 5459708
Conc: 1.66 ng/ml m



#11 alpha-Chlordane

R.T.: 4.974 min
Delta R.T.: -0.003 min
Response: 3612230
Conc: 1.07 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.973 min
Delta R.T.: 0.000 min
Response: 20168615
Conc: 11.81 ng/ml

Instrument :

ECD_L

ClientSampleId :


OR-02-062623

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/28/2023

Supervised By :Ankita Jodhani 06/28/2023



#28 Decachlorobiphenyl

R.T.: 7.881 min
Delta R.T.: 0.000 min
Response: 26901263
Conc: 10.88 ng/ml