

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062819\
 Data File : PL050040.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2019 18:49
 Operator : SM\AJ
 Sample : K3158-08MSD
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OK-02-062719MSD

Manual Integrations
 APPROVED

Ankita
 7/1/2019 10:19:17 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 23:12:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.321	3.933	182.0E6	56689720	16.140	18.515
2)	SA Decachlor...	7.985	8.964	220.7E6	63044760	17.366	18.780
Target Compounds							
2)	A alpha-BHC	3.749	4.321	851.1E6	227.5E6	45.302	50.809
3)	MA gamma-BHC...	4.024	4.604	811.2E6	212.9E6	45.754	51.318
4)	MA Heptachlor	4.309	5.106	841.4E6	216.5E6	46.288	50.008
5)	MB Aldrin	4.547	5.408	838.7E6	203.1E6	48.569	51.837
6)	B beta-BHC	4.279	4.780	347.2E6	96725682	47.463	49.935
7)	B delta-BHC	4.472	4.994	782.1E6	186.7E6	43.731	47.387
8)	B Heptachlo...	4.983	5.795	727.8E6	192.6E6	45.210	52.966
9)	A Endosulfan I	5.314	6.150	754.7E6	178.3E6	48.775	52.469
10)	B gamma-Chl...	5.205	6.029	816.0E6	193.1E6	47.748	51.821
11)	B alpha-Chl...	5.261	6.100	791.2E6	186.6E6	47.666	50.463
12)	B 4,4'-DDE	5.429	6.261	802.3E6	183.9E6	48.161	52.992
13)	MA Dieldrin	5.553	6.406	800.6E6	184.8E6	48.060	52.100
14)	MA Endrin	5.805	6.623	712.8E6	150.8E6	44.511	47.769
15)	B Endosulfa...	6.080	6.835	691.5E6	159.5E6	47.456	49.534
16)	A 4,4'-DDD	5.938	6.751	629.7E6	143.7E6	49.882m	54.954
17)	MA 4,4'-DDT	6.173	7.048	562.0E6	128.4E6	42.331	43.851
18)	B Endrin al...	6.246	6.959	558.5E6	135.9E6	47.285	49.467
19)	B Endosulfa...	6.458	7.182	617.1E6	151.7E6	47.595	50.411
20)	A Methoxychlor	6.718	7.509	261.6E6	70350072	40.130	42.207
21)	B Endrin ke...	6.939	7.650	705.3E6	169.3E6	50.182	54.330
22)	Mirex	7.113	8.099	506.5E6	133.3E6	48.999	52.702

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062819\
Data File : PL050040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2019 18:49
Operator : SM\AJ
Sample : K3158-08MSD
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
OK-02-062719MSD

Manual Integrations
APPROVED

Ankita
7/1/2019 10:19:17 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 23:12:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

