

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062824\
 Data File : PL090389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2024 17:23
 Operator : AR\AJ
 Sample : INDB301
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 04:13:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062824CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Jun 29 04:10:09 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.790	39229122	27528887	20.000	20.000
27)	SA Decachlor...	9.058	7.932	59989236	60319036	40.000	40.000
Target Compounds							
5)	MB Aldrin	5.264	4.245	48451985	33601255	20.000	20.000
6)	B beta-BHC	4.529	3.922	24803280	17586753	20.000	20.000
7)	B delta-BHC	4.776	4.152	50864748	35546266	20.000	20.000
8)	B Heptachlo...	5.690	4.747	47359236	33408605	20.000	20.000
10)	B trans-Chl...	5.945	4.997	44173804	32747928	20.000	20.000
11)	B cis-Chlor...	6.024	5.062	45081166	34202616	20.000	20.000
12)	B 4,4'-DDE	6.197	5.251	75385720	60979864	40.000	40.000
15)	B Endosulfa...	6.798	5.951	73476760	59025360	40.000	40.000
18)	B Endrin al...	6.928	6.131	57400950	46620442	40.000	40.000
19)	B Endosulfa...	7.162	6.353	68873685	58010072	40.000	40.000
21)	B Endrin ke...	7.646	6.859	74619898	66119112	40.000	40.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062824\
Data File : PL090389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2024 17:23
Operator : AR\AJ
Sample : INDB301
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 04:13:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062824CLP.M
Quant Title : GC Extractables
QLast Update : Sat Jun 29 04:10:09 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: 0.000 min
Response: 39229122
Conc: 20.00 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.790 min
Delta R.T.: 0.000 min
Response: 27528887
Conc: 20.00 ng/ml




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 48451985
Conc: 20.00 ng/ml




#5 Aldrin

R.T.: 4.245 min
Delta R.T.: 0.000 min
Response: 33601255
Conc: 20.00 ng/ml




#6 beta-BHC

R.T.: 4.529 min
Delta R.T.: 0.000 min
Response: 24803280
Conc: 20.00 ng/ml




#6 beta-BHC

R.T.: 3.922 min
Delta R.T.: 0.000 min
Response: 17586753
Conc: 20.00 ng/ml




#7 delta-BHC

R.T.: 4.776 min
Delta R.T.: 0.000 min
Response: 50864748
Conc: 20.00 ng/ml




#7 delta-BHC

R.T.: 4.152 min
Delta R.T.: 0.000 min
Response: 35546266
Conc: 20.00 ng/ml




#8 Heptachlor epoxide

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 47359236
Conc: 20.00 ng/ml




#8 Heptachlor epoxide

R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 33408605
Conc: 20.00 ng/ml




#10 trans-Chlordane

R.T.: 5.945 min
Delta R.T.: 0.000 min
Response: 44173804
Conc: 20.00 ng/ml




#10 trans-Chlordane

R.T.: 4.997 min
Delta R.T.: 0.000 min
Response: 32747928
Conc: 20.00 ng/ml




#11 cis-Chlordane

R.T.: 6.024 min
Delta R.T.: 0.000 min
Response: 45081166
Conc: 20.00 ng/ml




#11 cis-Chlordane

R.T.: 5.062 min
Delta R.T.: 0.000 min
Response: 34202616
Conc: 20.00 ng/ml




#12 4,4'-DDE

R.T.: 6.197 min
Delta R.T.: 0.000 min
Response: 75385720
Conc: 40.00 ng/ml




#12 4,4'-DDE

R.T.: 5.251 min
Delta R.T.: 0.000 min
Response: 60979864
Conc: 40.00 ng/ml




#15 Endosulfan II

R.T.: 6.798 min
Delta R.T.: 0.000 min
Response: 73476760
Conc: 40.00 ng/ml




#15 Endosulfan II

R.T.: 5.951 min
Delta R.T.: 0.000 min
Response: 59025360
Conc: 40.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.928 min
Delta R.T.: 0.000 min
Response: 57400950
Conc: 40.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.131 min
Delta R.T.: 0.000 min
Response: 46620442
Conc: 40.00 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.162 min
Delta R.T.: 0.000 min
Response: 68873685
Conc: 40.00 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.353 min
Delta R.T.: 0.000 min
Response: 58010072
Conc: 40.00 ng/ml




#21 Endrin ketone

R.T.: 7.646 min
Delta R.T.: 0.000 min
Response: 74619898
Conc: 40.00 ng/ml




#21 Endrin ketone

R.T.: 6.859 min
Delta R.T.: 0.000 min
Response: 66119112
Conc: 40.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 59989236
Conc: 40.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 60319036
Conc: 40.00 ng/ml