

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062921\
 Data File : PL067865.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2021 08:53
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 02:55:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 11 02:00:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	5.529	11318569	18695453	20.587	19.831
28)	SA Decachlor...	10.219	11.427	14937021	20172544	21.077	20.252
Target Compounds							
4)	MA Heptachlor	5.987f	0.000	539492	0	0.715	N.D. #
6)	B beta-BHC	5.987	0.000	539492	0	1.508	N.D. #
15)	B Endosulfa...	0.000	9.180f	0	413609	N.D.	0.519 #
24)	Chlordane-2	6.485	0.000	287363	0	10.731	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL062921\
Data File : PL067865.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2021 08:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 02:55:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL061021.M
Quant Title : GC Extractables
QLast Update : Fri Jun 11 02:00:12 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: -0.003 min
Response: 11318569
Conc: 20.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 5.529 min
Delta R.T.: -0.002 min
Response: 18695453
Conc: 19.83 ng/ml




#4 Heptachlor

R.T.: 5.987 min
Delta R.T.: -0.025 min
Response: 539492
Conc: 0.72 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 7.155 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 5.987 min
Delta R.T.: 0.004 min
Response: 539492
Conc: 1.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 6.762 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.190 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 9.180 min
Delta R.T.: 0.021 min
Response: 413609
Conc: 0.52 ng/ml



#24 Chlordane-2


R.T.: 6.485 min
Delta R.T.: -0.002 min
Response: 287363
Conc: 10.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 7.503 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.219 min
Delta R.T.: -0.006 min
Response: 14937021
Conc: 21.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.427 min
Delta R.T.: -0.002 min
Response: 20172544
Conc: 20.25 ng/ml