

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL063020\
 Data File : PL059560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2020 11:23
 Operator : AJ\MA
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 12:26:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL063020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 30 12:24:14 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.517	4.071	178.7E6	264.4E6	98.687	97.699
2)	SA Decachlor...	8.287	9.182	148.5E6	204.6E6	95.867	93.414
Target Compounds							
2)	A alpha-BHC	3.958	4.463	269.3E6	411.3E6	102.547	101.372
3)	MA gamma-BHC...	4.243	4.751	236.2E6	381.6E6	101.581	100.736
4)	MA Heptachlor	4.548	5.265	209.6E6	352.2E6	98.437	98.506
5)	MB Aldrin	4.798	5.572	171.4E6	324.7E6	97.556	100.385
6)	B beta-BHC	4.492	4.919	89426209	152.1E6	94.993	96.231
7)	B delta-BHC	4.699	5.139	198.7E6	304.4E6	97.319	99.792
8)	B Heptachlo...	5.242	5.960	161.3E6	260.8E6	98.272	99.937
9)	A Endosulfan I	5.582	6.319	153.6E6	268.8E6	98.726	97.662
10)	B gamma-Chl...	5.469	6.195	164.5E6	295.7E6	99.294	98.278
11)	B alpha-Chl...	5.528	6.269	164.1E6	294.3E6	98.701	97.721
12)	B 4,4'-DDE	5.691	6.423	157.1E6	278.3E6	102.036	100.860
13)	MA Dieldrin	5.826	6.578	164.5E6	286.6E6	100.426	99.456
14)	MA Endrin	6.083	6.797	151.0E6	248.0E6	99.867	99.083
15)	B Endosulfa...	6.356	7.005	148.6E6	247.7E6	97.229	96.395
16)	A 4,4'-DDD	6.208	6.915	134.2E6	233.3E6	99.729	98.094
17)	MA 4,4'-DDT	6.449	7.215	137.2E6	232.2E6	102.688	99.756
18)	B Endrin al...	6.524	7.129	123.8E6	211.9E6	96.600	95.705
19)	B Endosulfa...	6.737	7.352	141.1E6	230.1E6	100.106	94.777
20)	A Methoxychlor	6.994	7.672	75964382	123.1E6	96.524	93.771
21)	B Endrin ke...	7.229	7.826	165.5E6	246.8E6	97.674	95.782
22)	Mirex	7.422	8.283	124.2E6	176.5E6	96.674	93.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL063020\
Data File : PL059560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2020 11:23
Operator : AJ\MA
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 12:26:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL063020.M
Quant Title : GC Extractables
QLast Update : Tue Jun 30 12:24:14 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

