

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070119\
 Data File : PL050094.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2019 18:27
 Operator : SM\AJ
 Sample : K3560-18
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 EXT-012-SW139-062419

Manual Integrations
 APPROVED

Ankita
 7/2/2019 1:10:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 01:29:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.320	3.933	221.9E6	57309721	19.680	18.718
28)	SA Decachlor...	7.986	8.966	204.3E6	54525179	16.075	16.242
Target Compounds							
17)	MA 4,4'-DDT	6.171	7.049	9960893	1733857	0.750m	0.592m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070119\
Data File : PL050094.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2019 18:27
Operator : SM\AJ
Sample : K3560-18
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
EXT-012-SW139-062419


Manual Integrations
APPROVED

Ankita
7/2/2019 1:10:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 01:29:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.320 min
Delta R.T.: -0.004 min
Response: 221949973
Conc: 19.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
EXT-012-SW139-062419


Manual Integrations
APPROVED

Ankita
7/2/2019 1:10:22 PM




#1 Tetrachloro-m-xylene

R.T.: 3.933 min
Delta R.T.: -0.002 min
Response: 57309721
Conc: 18.72 ng/ml




#17 4,4'-DDT

R.T.: 6.171 min
Delta R.T.: -0.009 min
Response: 9960893
Conc: 0.75 ng/ml m



#17 4,4'-DDT

R.T.: 7.049 min
Delta R.T.: -0.005 min
Response: 1733857
Conc: 0.59 ng/ml m




#28 Decachlorobiphenyl

R.T.: 7.986 min
Delta R.T.: -0.006 min
Response: 204250750
Conc: 16.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
EXT-012-SW139-062419

Manual Integrations
APPROVED

Ankita
7/2/2019 1:10:22 PM



#28 Decachlorobiphenyl

R.T.: 8.966 min
Delta R.T.: -0.007 min
Response: 54525179
Conc: 16.24 ng/ml