

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070124\
 Data File : PL090420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2024 16:48
 Operator : AR\AJ
 Sample : P3135-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OK-628-COMP-02

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/02/2024
 Supervised By :Ankita Jodhani 07/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 01:52:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062424.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 25 11:05:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.546	2.788	30403866	21829050	13.396	13.801m
28) SA Decachlor...	9.056	7.931	21611271	20075349	14.013	13.458m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070124\
Data File : PL090420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2024 16:48
Operator : AR\AJ
Sample : P3135-07
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
OK-628-COMP-02


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/02/2024
Supervised By : Ankita Jodhani 07/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 01:52:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062424.M
Quant Title : GC Extractables
QLast Update : Tue Jun 25 11:05:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: -0.001 min
Response: 30403866
Conc: 13.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
OK-628-COMP-02


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/02/2024
Supervised By :Ankita Jodhani 07/02/2024




#1 Tetrachloro-m-xylene

R.T.: 2.788 min
Delta R.T.: -0.002 min
Response: 21829050
Conc: 13.80 ng/ml m



#28 Decachlorobiphenyl

R.T.: 9.056 min
Delta R.T.: -0.005 min
Response: 21611271
Conc: 14.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.931 min
Delta R.T.: -0.005 min
Response: 20075349
Conc: 13.46 ng/ml m