

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070318\
 Data File : PL037653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jul 2018 12:19
 Operator : UA\AJ
 Sample : IDOC WATER 01
 Misc : FOR SJ
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 IDOC WATER 01

Manual Integrations
 APPROVED

mohammad
 7/5/2018 2:13:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 04 00:10:33 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 21 02:27:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.337	3.935	115.2E6	44386256	14.561	16.043
28)	SA Decachlor...	8.014	8.946	92698758	28789489	15.363	14.511
Target Compounds							
2)	A alpha-BHC	3.765	4.314	517.1E6	183.7E6	42.237m	47.348
3)	MA gamma-BHC...	4.044	4.594	455.3E6	162.9E6	44.144	47.158
4)	MA Heptachlor	4.328	5.105	452.8E6	159.8E6	44.192	46.234
5)	MB Aldrin	4.568	5.407	442.8E6	145.2E6	43.501	46.274
6)	B beta-BHC	4.298	4.756	200.3E6	72288816	43.923	44.825
7)	B delta-BHC	4.493	4.970	460.9E6	148.9E6	43.567	52.808
8)	B Heptachlo...	5.007	5.787	413.4E6	132.4E6	44.130	46.431
9)	A Endosulfan I	5.337	6.140	371.0E6	116.1E6	43.851	45.263
10)	B gamma-Chl...	5.229	6.020	415.7E6	130.7E6	44.386	46.394
11)	B alpha-Chl...	5.285	6.093	401.7E6	130.5E6	44.229	45.325
12)	B 4,4'-DDE	5.454	6.250	397.2E6	116.2E6	44.971	45.673
13)	MA Dieldrin	5.577	6.394	397.9E6	121.7E6	44.949	45.597
14)	MA Endrin	5.831	6.608	348.5E6	103.5E6	44.240	45.010
15)	B Endosulfa...	6.106	6.811	331.4E6	107.0E6	45.064	43.138
16)	A 4,4'-DDD	5.966	6.731	307.1E6	92404484	46.771	45.016
17)	MA 4,4'-DDT	6.201	7.030	288.8E6	91954755	44.730	43.412
18)	B Endrin al...	6.273	6.934	258.0E6	88769795	46.195	44.114
19)	B Endosulfa...	6.484	7.156	297.0E6	96696824	44.394	44.429
20)	A Methoxychlor	6.749	7.485	147.1E6	51603170	43.280	46.101
21)	B Endrin ke...	6.968	7.620	304.5E6	98682998	42.816	48.620

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070318\
Data File : PL037653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jul 2018 12:19
Operator : UA\AJ
Sample : IDOC WATER 01
Misc : FOR SJ
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
IDOC WATER 01

Manual Integrations
APPROVED

mohammad
7/5/2018 2:13:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 04 00:10:33 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
Quant Title : GC Extractables
QLast Update : Thu Jun 21 02:27:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

