

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070318\
 Data File : PL037659.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jul 2018 14:11
 Operator : UA\AJ
 Sample : PSTDCCC050
 Misc : FOR SJ
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 7/5/2018 2:14:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 04 00:11:02 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 21 02:27:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.333	3.935	337.4E6	132.2E6	42.658	47.786
28)	SA Decachlor...	8.011	8.945	285.0E6	90562853	47.227	45.648
Target Compounds							
2)	A alpha-BHC	3.763	4.314	534.4E6	193.4E6	43.646	49.830
3)	MA gamma-BHC...	4.040	4.594	464.2E6	170.7E6	45.001	49.425
4)	MA Heptachlor	4.325	5.105	472.1E6	167.6E6	46.079	48.503
5)	MB Aldrin	4.564	5.408	465.1E6	153.0E6	45.696	48.770
6)	B beta-BHC	4.295	4.755	210.4E6	76072269	46.129	47.171
7)	B delta-BHC	4.489	4.970	478.3E6	158.7E6	45.215	56.278
8)	B Heptachlo...	5.003	5.788	431.5E6	139.8E6	46.072	49.034
9)	A Endosulfan I	5.334	6.140	391.9E6	122.2E6	46.323	47.661
10)	B gamma-Chl...	5.225	6.020	434.9E6	139.9E6	46.434	49.672
11)	B alpha-Chl...	5.282	6.094	419.2E6	136.8E6	46.159	47.541
12)	B 4,4'-DDE	5.451	6.249	414.9E6	122.9E6	46.979	48.278
13)	MA Dieldrin	5.574	6.394	412.1E6	127.9E6	46.546	47.902
14)	MA Endrin	5.828	6.608	358.1E6	109.3E6	45.457	47.555
15)	B Endosulfa...	6.102	6.811	344.8E6	112.1E6	46.894	45.220
16)	A 4,4'-DDD	5.963	6.730	311.4E6	97550130	47.412	47.523
17)	MA 4,4'-DDT	6.197	7.029	300.2E6	97364390	46.503	45.966
18)	B Endrin al...	6.270	6.933	262.9E6	91721895	47.084	45.581
19)	B Endosulfa...	6.481	7.156	306.2E6	101.5E6	45.775	46.639
20)	A Methoxychlor	6.744	7.485	155.7E6	54314272	45.803m	48.523
21)	B Endrin ke...	6.963	7.620	333.3E6	105.3E6	46.861m	51.891

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070318\
Data File : PL037659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jul 2018 14:11
Operator : UA\AJ
Sample : PSTDCCC050
Misc : FOR SJ
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
7/5/2018 2:14:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 04 00:11:02 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
Quant Title : GC Extractables
QLast Update : Thu Jun 21 02:27:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

