

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070518\
 Data File : PL037696.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2018 14:43
 Operator : UA\AJ
 Sample : J3853-12
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 14:57:49 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 21 02:27:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.333	3.935	132.7E6	52030869	16.778	18.806
28)	SA Decachlor...	8.010	8.944	81045226	27313676	13.431	13.767
Target Compounds							
2)	A alpha-BHC	3.718f	0.000	3958510	0	0.323	N.D. #
5)	MB Aldrin	4.553	0.000	46783729	0	4.596	N.D. #
7)	B delta-BHC	4.448f	0.000	5098270	0	0.482	N.D. #
8)	B Heptachlo...	4.985	0.000	34589740	0	3.693	N.D. #
10)	B gamma-Chl...	5.225	0.000	4580959	0	0.489	N.D. #
11)	B alpha-Chl...	5.225f	0.000	4580959	0	0.504	N.D. #
16)	A 4,4'-DDD	5.962	0.000	23291417	0	3.472	N.D. #
25)	Chlordane-3	5.225	0.000	4580959	0	4.310	N.D. #
26)	Chlordane-4	5.225	0.000	4580959	0	4.754	N.D. #
27)	Chlordane-5	5.962	0.000	23291417	0	179.676	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070518\
Data File : PL037696.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2018 14:43
Operator : UA\AJ
Sample : J3853-12
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 14:57:49 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
Quant Title : GC Extractables
QLast Update : Thu Jun 21 02:27:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.333 min
Delta R.T.: -0.004 min
Response: 132716628
Conc: 16.78 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.935 min
Delta R.T.: -0.003 min
Response: 52030869
Conc: 18.81 ng/ml




#2 alpha-BHC

R.T.: 3.718 min
Delta R.T.: -0.050 min
Response: 3958510
Conc: 0.32 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 4.317 min
Response: 0
Conc: N.D.



#5 Aldrin


R.T.: 4.553 min
Delta R.T.: -0.017 min
Response: 46783729
Conc: 4.60 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 5.411 min
Response: 0
Conc: N.D.




#7 delta-BHC

R.T.: 4.448 min
Delta R.T.: -0.047 min
Response: 5098270
Conc: 0.48 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 4.972 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 4.985 min
Delta R.T.: -0.025 min
Response: 34589740
Conc: 3.69 ng/ml

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.790 min
Response: 0
Conc: N.D.




#10 gamma-Chlordane

R.T.: 5.225 min
Delta R.T.: -0.007 min
Response: 4580959
Conc: 0.49 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 6.023 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 5.225 min
Delta R.T.: -0.064 min
Response: 4580959
Conc: 0.50 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 6.097 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 5.962 min
Delta R.T.: -0.008 min
Response: 23291417
Conc: 3.47 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T. : 6.734 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 5.225 min
Delta R.T.: -0.008 min
Response: 4580959
Conc: 4.31 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 6.023 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 5.225 min
Delta R.T.: -0.064 min
Response: 4580959
Conc: 4.75 ng/ml



#26 Chlordane-4


R.T.: 0.000 min
Exp R.T. : 6.099 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.962 min
Delta R.T.: 0.064 min
Response: 23291417
Conc: 179.68 ng/ml

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.890 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.010 min
Delta R.T.: -0.011 min
Response: 81045226
Conc: 13.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.944 min
Delta R.T.: -0.007 min
Response: 27313676
Conc: 13.77 ng/ml