

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070519\
 Data File : PL050197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2019 10:28
 Operator : SM\AJ
 Sample : K3624-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 23:05:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	3.932	148.1E6	36439160	13.131	11.901
28)	SA Decachlor...	7.983	8.963	110.0E6	32954345	8.657	9.816
Target Compounds							
10)	B gamma-Chl...	0.000	6.027	0	9051836	N.D.	2.429 #
11)	B alpha-Chl...	0.000	6.100	0	11771692	N.D.	3.183 #
17)	MA 4,4'-DDT	6.171	0.000	30073967	0	2.265	N.D. #
24)	Chlordane-2	4.683f	0.000	19674805	0	33.688	N.D. #
25)	Chlordane-3	0.000	6.027	0	9051836	N.D.	18.190 #
26)	Chlordane-4	0.000	6.100	0	11771692	N.D.	19.806 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070519\
Data File : PL050197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2019 10:28
Operator : SM\AJ
Sample : K3624-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 23:05:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.322 min
Delta R.T.: -0.002 min
Response: 148087859
Conc: 13.13 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.932 min
Delta R.T.: -0.003 min
Response: 36439160
Conc: 11.90 ng/ml




#10 gamma-Chlordane

R.T.: 0.000 min
Exp R.T. : 5.210 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 6.027 min
Delta R.T.: -0.007 min
Response: 9051836
Conc: 2.43 ng/ml



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.266 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 6.100 min
Delta R.T.: -0.006 min
Response: 11771692
Conc: 3.18 ng/ml




#17 4,4'-DDT

R.T.: 6.171 min
Delta R.T.: -0.009 min
Response: 30073967
Conc: 2.27 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 7.054 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 4.683 min
Delta R.T.: 0.028 min
Response: 19674805
Conc: 33.69 ng/ml

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 5.390 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 5.211 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 6.027 min
Delta R.T.: -0.007 min
Response: 9051836
Conc: 18.19 ng/ml



#26 Chlordane-4


R.T.: 0.000 min
Exp R.T.: 5.266 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 6.100 min
Delta R.T.: -0.008 min
Response: 11771692
Conc: 19.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: -0.009 min
Response: 110002577
Conc: 8.66 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.963 min
Delta R.T.: -0.011 min
Response: 32954345
Conc: 9.82 ng/ml