

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070522\
 Data File : PL076247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2022 13:03
 Operator : AR\AJ
 Sample : N3564-31 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WC-16-2-3

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/06/2022
 Supervised By : Ankita Jodhani 07/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 06 02:53:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.636	4.624	1010754	2434360	1.336m	1.284m
2) SA Decachlor...	11.623	10.361	1875975	4086892	2.981	2.661

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070522\
Data File : PL076247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2022 13:03
Operator : AR\AJ
Sample : N3564-31 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
WC-16-2-3


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/06/2022
Supervised By : Ankita Jodhani 07/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 06 02:53:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.636 min
Delta R.T.: -0.004 min
Response: 1010754
Conc: 1.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
WC-16-2-3


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/06/2022
Supervised By :Ankita Jodhani 07/06/2022




#1 Tetrachloro-m-xylene

R.T.: 4.624 min
Delta R.T.: -0.009 min
Response: 2434360
Conc: 1.28 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.623 min
Delta R.T.: -0.010 min
Response: 1875975
Conc: 2.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.361 min
Delta R.T.: -0.009 min
Response: 4086892
Conc: 2.66 ng/ml