

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070524\
 Data File : PL090443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2024 14:22
 Operator : AR\AJ
 Sample : P3166-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 NB-08-070324

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 06 01:09:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062424.M
 Quant Title : GC Extractables
 QLast Update : Tue Jun 25 11:05:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.788	20717276	16505043	9.128	10.435m
28)	SA Decachlor...	9.059	7.933	17846888	15448081	11.572	10.356
Target Compounds							
10)	B gamma-Chl...	5.945	5.000	7118979	1617778	2.817m	0.844m#
11)	B alpha-Chl...	6.028	5.061	17577647	4893307	7.047m	2.555 #
12)	B 4,4'-DDE	6.196	5.250	8431297	6088390	3.862m	3.675
17)	MA 4,4'-DDT	7.026	6.054	7581476	6052567	3.951m	4.089



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070524\
Data File : PL090443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2024 14:22
Operator : AR\AJ
Sample : P3166-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
NB-08-070324

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 06 01:09:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062424.M
Quant Title : GC Extractables
QLast Update : Tue Jun 25 11:05:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.546 min
Delta R.T.: 0.000 min
Response: 20717276
Conc: 9.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
NB-08-070324




#1 Tetrachloro-m-xylene

R.T.: 2.788 min
Delta R.T.: -0.002 min
Response: 16505043
Conc: 10.43 ng/ml m




#10 gamma-Chlordane

R.T.: 5.945 min
Delta R.T.: -0.003 min
Response: 7118979
Conc: 2.82 ng/ml m



#10 gamma-Chlordane


R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 1617778
Conc: 0.84 ng/ml m



#11 alpha-Chlordane


R.T.: 6.028 min
Delta R.T.: 0.001 min
Response: 17577647
Conc: 7.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
NB-08-070324




#11 alpha-Chlordane

R.T.: 5.061 min
Delta R.T.: -0.003 min
Response: 4893307
Conc: 2.55 ng/ml




#12 4,4'-DDE

R.T.: 6.196 min
Delta R.T.: -0.004 min
Response: 8431297
Conc: 3.86 ng/ml m



#12 4,4'-DDE


R.T.: 5.250 min
Delta R.T.: -0.002 min
Response: 6088390
Conc: 3.68 ng/ml



#17 4,4'-DDT


R.T.: 7.026 min
Delta R.T.: -0.004 min
Response: 7581476
Conc: 3.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
NB-08-070324




#17 4,4'-DDT

R.T.: 6.054 min
Delta R.T.: -0.003 min
Response: 6052567
Conc: 4.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.059 min
Delta R.T.: -0.002 min
Response: 17846888
Conc: 11.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.933 min
Delta R.T.: -0.004 min
Response: 15448081
Conc: 10.36 ng/ml