

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070622\
 Data File : PL076262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jul 2022 10:34
 Operator : AR\AJ
 Sample : PB146012BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB146012BL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/07/2022
 Supervised By : Ankita Jodhani 07/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 00:07:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.638	4.626	11888925	32332870	15.719	17.048
28) SA Decachlor...	11.623	10.361	8679370	33615993	13.793m	21.887 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070622\
Data File : PL076262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jul 2022 10:34
Operator : AR\AJ
Sample : PB146012BL
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB146012BL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/07/2022
Supervised By : Ankita Jodhani 07/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 00:07:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.638 min
Delta R.T.: -0.002 min
Response: 11888925
Conc: 15.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB146012BL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/07/2022
Supervised By :Ankita Jodhani 07/07/2022




#1 Tetrachloro-m-xylene

R.T.: 4.626 min
Delta R.T.: -0.008 min
Response: 32332870
Conc: 17.05 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.623 min
Delta R.T.: -0.011 min
Response: 8679370
Conc: 13.79 ng/ml m



#28 Decachlorobiphenyl

R.T.: 10.361 min
Delta R.T.: -0.009 min
Response: 33615993
Conc: 21.89 ng/ml