

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070722\
 Data File : PL076327.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2022 22:58
 Operator : AR\AJ
 Sample : N3606-01
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 QUARTERLY-INFLUENT

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/08/2022
 Supervised By :Ankita Jodhani 07/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 04:46:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 24 02:17:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.636	4.623	7002068	16536348	9.258m	8.719m
28) SA Decachlor...	11.650f	10.357	23515670	4700180	37.370	3.060m#
Target Compounds						
16) A 4,4'-DDD	9.194	8.193	3850384	8424517	6.790	6.074

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070722\
Data File : PL076327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2022 22:58
Operator : AR\AJ
Sample : N3606-01
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
QUARTERLY-INFLUENT


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/08/2022
Supervised By :Ankita Jodhani 07/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 04:46:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062222.M
Quant Title : GC Extractables
QLast Update : Fri Jun 24 02:17:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.636 min
Delta R.T.: -0.004 min
Response: 7002068
Conc: 9.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
QUARTERLY-INFLUENT


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/08/2022
Supervised By :Ankita Jodhani 07/08/2022




#1 Tetrachloro-m-xylene

R.T.: 4.623 min
Delta R.T.: -0.010 min
Response: 16536348
Conc: 8.72 ng/ml m




#16 4,4'-DDD

R.T.: 9.194 min
Delta R.T.: -0.004 min
Response: 3850384
Conc: 6.79 ng/ml



#16 4,4'-DDD

R.T.: 8.193 min
Delta R.T.: -0.008 min
Response: 8424517
Conc: 6.07 ng/ml




#28 Decachlorobiphenyl

R.T.: 11.650 min
Delta R.T.: 0.017 min
Response: 23515670
Conc: 37.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
QUARTERLY-INFLUENT

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/08/2022
Supervised By :Ankita Jodhani 07/08/2022



#28 Decachlorobiphenyl

R.T.: 10.357 min
Delta R.T.: -0.013 min
Response: 4700180
Conc: 3.06 ng/ml m