

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070819\
 Data File : PL050210.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2019 11:06
 Operator : SM\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 03:28:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds									
14)	MA	Endrin	5.781f	0.000	260.9E6	0	16.290	N.D.	#
17)	MA	4,4'-DDT	6.162f	0.000	382.6E6	0	28.817	N.D.	#
20)	A	Methoxychlor	6.733	0.000	54192458	0	8.314	N.D.	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070819\
Data File : PL050210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2019 11:06
Operator : SM\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 03:28:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T.: 3.324 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.935 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 5.781 min
Delta R.T.: -0.029 min
Response: 260882729
Conc: 16.29 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.629 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 6.162 min
Delta R.T.: -0.018 min
Response: 382566385
Conc: 28.82 ng/ml

Instrument :
ECD_L
ClientSampleId :
HEXANE




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 7.054 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 6.733 min
Delta R.T.: 0.008 min
Response: 54192458
Conc: 8.31 ng/ml



#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 7.515 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl


R.T.: 0.000 min

Exp R.T. : 7.992 min

Response: 0

Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
HEXANE



#28 Decachlorobiphenyl

R.T.: 0.000 min

Exp R.T. : 8.973 min

Response: 0

Conc: N.D.