

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070819\
 Data File : PL050216.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2019 13:31
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/9/2019 9:31:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 03:30:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.321	3.932	640.4E6	164.4E6	56.786	53.710
28)	SA Decachlor...	7.985	8.964	567.7E6	150.3E6	44.682	44.784
Target Compounds							
2)	A alpha-BHC	3.749	4.321	1082.9E6	243.9E6	57.642	54.482
3)	MA gamma-BHC...	4.024	4.604	1001.7E6	229.1E6	56.498	55.226
4)	MA Heptachlor	4.308	5.106	1053.4E6	242.7E6	57.952	56.060
5)	MB Aldrin	4.546	5.408	1012.2E6	224.1E6	58.619	57.213
6)	B beta-BHC	4.278	4.779	417.3E6	104.1E6	57.044	53.732
7)	B delta-BHC	4.472	4.993	1045.3E6	222.6E6	58.447	56.511
8)	B Heptachlor...	4.983	5.795	917.4E6	209.1E6	56.991	57.522
9)	A Endosulfan I	5.314	6.150	847.0E6	188.9E6	54.745	55.595
10)	B gamma-Chl...	5.205	6.029	968.8E6	212.3E6	56.692	56.979
11)	B alpha-Chl...	5.261	6.101	935.3E6	208.2E6	56.344	56.303
12)	B 4,4'-DDE	5.429	6.261	813.5E6	203.7E6	48.831	58.702
13)	MA Dieldrin	5.552	6.406	936.1E6	200.9E6	56.198	56.619
14)	MA Endrin	5.805	6.624	863.3E6	169.9E6	53.906	53.816
15)	B Endosulfa...	6.080	6.835	799.5E6	171.0E6	54.871	53.092
16)	A 4,4'-DDD	5.938	6.751	747.9E6	156.4E6	59.246m	59.842
17)	MA 4,4'-DDT	6.173	7.048	683.5E6	143.3E6	51.486	48.941
18)	B Endrin al...	6.246	6.959	633.5E6	144.6E6	53.635	52.645
19)	B Endosulfa...	6.458	7.182	677.7E6	157.2E6	52.266	52.246
20)	A Methoxychlor	6.718	7.508	330.4E6	79558675	50.691	47.732
21)	B Endrin ke...	6.939	7.651	724.5E6	160.3E6	51.543	51.451
22)	Mirex	7.113	8.100	494.0E6	121.5E6	47.790	48.060

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070819\
Data File : PL050216.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2019 13:31
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
7/9/2019 9:31:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 03:30:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

