

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070825\
 Data File : PL096262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 11:58
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/09/2025
 Supervised By :mohammad ahmed 07/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 01:56:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 07 15:22:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.826	167.1E6	343.7E6	45.804	59.557 #
28)	SA Decachlor...	9.022	7.997	127.4E6	231.4E6	45.288	46.403
Target Compounds							
23)	Chlordane-1	4.691	3.840	86803615	114.5E6	484.027m	484.231
24)	Chlordane-2	5.217	4.420	93928942	131.8E6	479.768	479.540
25)	Chlordane-3	5.920	5.054	355.8E6	388.1E6	482.903	482.443
26)	Chlordane-4	6.005	5.118	431.0E6	344.9E6	480.383	458.927
27)	Chlordane-5	6.842	6.012	69880955	135.9E6	462.982	479.508

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL070825\
Data File : PL096262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 11:58
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/09/2025
Supervised By :mohammad ahmed 07/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 01:56:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070725.M
Quant Title : GC Extractables
QLast Update : Mon Jul 07 15:22:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

