

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071019\
 Data File : PL050285.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2019 18:10
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleID :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/11/2019 8:18:09 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 01:21:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	3.933	584.6E6	167.0E6	51.838	54.541
28)	SA Decachlor...	7.984	8.965	559.7E6	151.3E6	44.048	45.057
Target Compounds							
2)	A alpha-BHC	3.749	4.321	999.5E6	251.1E6	53.200	56.099
3)	MA gamma-BHC...	4.025	4.604	924.6E6	233.8E6	52.146	56.362
4)	MA Heptachlor	4.308	5.106	968.8E6	245.1E6	53.300	56.634
5)	MB Aldrin	4.547	5.408	917.0E6	226.9E6	53.103	57.923
6)	B beta-BHC	4.279	4.780	382.7E6	107.8E6	52.313	55.661
7)	B delta-BHC	4.473	4.994	954.7E6	229.5E6	53.381	58.254
8)	B Heptachlor...	4.983	5.795	843.2E6	212.9E6	52.385	58.560
9)	A Endosulfan I	5.314	6.150	798.6E6	192.3E6	51.617	56.607
10)	B gamma-Chl...	5.205	6.029	897.2E6	215.7E6	52.505	57.879
11)	B alpha-Chl...	5.261	6.101	867.6E6	211.7E6	52.271	57.230
12)	B 4,4'-DDE	5.430	6.262	759.3E6	207.3E6	45.576	59.725 #
13)	MA Dieldrin	5.553	6.407	876.4E6	204.4E6	52.611	57.617
14)	MA Endrin	5.805	6.624	826.1E6	171.1E6	51.583	54.208
15)	B Endosulfa...	6.080	6.836	755.9E6	174.7E6	51.875	54.236
16)	A 4,4'-DDD	5.939	6.750	692.0E6	156.7E6	54.820m	59.931m
17)	MA 4,4'-DDT	6.174	7.049	639.3E6	144.3E6	48.154	49.255
18)	B Endrin al...	6.247	6.960	603.0E6	146.9E6	51.053	53.498
19)	B Endosulfa...	6.458	7.183	651.4E6	160.3E6	50.244	53.265
20)	A Methoxychlor	6.719	7.509	303.6E6	79614815	46.578	47.766
21)	B Endrin ke...	6.940	7.651	695.7E6	166.1E6	49.499	53.317
22)	Mirex	7.113	8.100	477.2E6	124.0E6	46.164	49.056

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071019\
Data File : PL050285.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2019 18:10
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 01:21:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.322 min
Delta R.T.: -0.002 min
Response: 584617848
Conc: 51.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#1 Tetrachloro-m-xylene

R.T.: 3.933 min
Delta R.T.: -0.003 min
Response: 166991025
Conc: 54.54 ng/ml




#2 alpha-BHC

R.T.: 3.749 min
Delta R.T.: -0.002 min
Response: 999491835
Conc: 53.20 ng/ml



#2 alpha-BHC

R.T.: 4.321 min
Delta R.T.: -0.003 min
Response: 251132752
Conc: 56.10 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.025 min
Delta R.T.: -0.003 min
Response: 924577704
Conc: 52.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#3 gamma-BHC (Lindane)

R.T.: 4.604 min
Delta R.T.: -0.003 min
Response: 233778597
Conc: 56.36 ng/ml




#4 Heptachlor

R.T.: 4.308 min
Delta R.T.: -0.004 min
Response: 968842040
Conc: 53.30 ng/ml



#4 Heptachlor

R.T.: 5.106 min
Delta R.T.: -0.004 min
Response: 245134574
Conc: 56.63 ng/ml




#5 Aldrin

R.T.: 4.547 min
Delta R.T.: -0.004 min
Response: 916972391
Conc: 53.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#5 Aldrin

R.T.: 5.408 min
Delta R.T.: -0.005 min
Response: 226886978
Conc: 57.92 ng/ml




#6 beta-BHC

R.T.: 4.279 min
Delta R.T.: -0.003 min
Response: 382661759
Conc: 52.31 ng/ml



#6 beta-BHC

R.T.: 4.780 min
Delta R.T.: -0.003 min
Response: 107817888
Conc: 55.66 ng/ml




#7 delta-BHC

R.T.: 4.473 min
Delta R.T.: -0.004 min
Response: 954650339
Conc: 53.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#7 delta-BHC

R.T.: 4.994 min
Delta R.T.: -0.003 min
Response: 229454456
Conc: 58.25 ng/ml




#8 Heptachlor epoxide

R.T.: 4.983 min
Delta R.T.: -0.005 min
Response: 843248577
Conc: 52.38 ng/ml



#8 Heptachlor epoxide

R.T.: 5.795 min
Delta R.T.: -0.005 min
Response: 212911757
Conc: 58.56 ng/ml




#9 Endosulfan I

R.T.: 5.314 min
Delta R.T.: -0.005 min
Response: 798639740
Conc: 51.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#9 Endosulfan I

R.T.: 6.150 min
Delta R.T.: -0.005 min
Response: 192315231
Conc: 56.61 ng/ml




#10 gamma-Chlordane

R.T.: 5.205 min
Delta R.T.: -0.005 min
Response: 897245768
Conc: 52.50 ng/ml



#10 gamma-Chlordane

R.T.: 6.029 min
Delta R.T.: -0.005 min
Response: 215684682
Conc: 57.88 ng/ml




#11 alpha-Chlordane

R.T.: 5.261 min
Delta R.T.: -0.005 min
Response: 867630308
Conc: 52.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#11 alpha-Chlordane

R.T.: 6.101 min
Delta R.T.: -0.005 min
Response: 211661750
Conc: 57.23 ng/ml




#12 4,4'-DDE

R.T.: 5.430 min
Delta R.T.: -0.005 min
Response: 759259947
Conc: 45.58 ng/ml



#12 4,4'-DDE

R.T.: 6.262 min
Delta R.T.: -0.004 min
Response: 207252070
Conc: 59.72 ng/ml




#13 Dielldrin

R.T.: 5.553 min
Delta R.T.: -0.005 min
Response: 876382205
Conc: 52.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#13 Dielldrin

R.T.: 6.407 min
Delta R.T.: -0.005 min
Response: 204399316
Conc: 57.62 ng/ml




#14 Endrin

R.T.: 5.805 min
Delta R.T.: -0.005 min
Response: 826080387
Conc: 51.58 ng/ml



#14 Endrin

R.T.: 6.624 min
Delta R.T.: -0.005 min
Response: 171146700
Conc: 54.21 ng/ml




#15 Endosulfan II

R.T.: 6.080 min
Delta R.T.: -0.005 min
Response: 755882745
Conc: 51.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#15 Endosulfan II

R.T.: 6.836 min
Delta R.T.: -0.005 min
Response: 174689816
Conc: 54.24 ng/ml




#16 4,4'-DDD

R.T.: 5.939 min
Delta R.T.: -0.006 min
Response: 692043914
Conc: 54.82 ng/ml m



#16 4,4'-DDD

R.T.: 6.750 min
Delta R.T.: -0.007 min
Response: 156674053
Conc: 59.93 ng/ml m




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: -0.006 min
Response: 639276725
Conc: 48.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#17 4,4'-DDT

R.T.: 7.049 min
Delta R.T.: -0.005 min
Response: 144256950
Conc: 49.25 ng/ml




#18 Endrin aldehyde

R.T.: 6.247 min
Delta R.T.: -0.005 min
Response: 603025850
Conc: 51.05 ng/ml



#18 Endrin aldehyde

R.T.: 6.960 min
Delta R.T.: -0.005 min
Response: 146926108
Conc: 53.50 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.458 min
Delta R.T.: -0.006 min
Response: 651434335
Conc: 50.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#19 Endosulfan Sulfate

R.T.: 7.183 min
Delta R.T.: -0.005 min
Response: 160305899
Conc: 53.27 ng/ml




#20 Methoxychlor

R.T.: 6.719 min
Delta R.T.: -0.006 min
Response: 303606998
Conc: 46.58 ng/ml



#20 Methoxychlor

R.T.: 7.509 min
Delta R.T.: -0.005 min
Response: 79614815
Conc: 47.77 ng/ml




#21 Endrin ketone

R.T.: 6.940 min
Delta R.T.: -0.006 min
Response: 695719862
Conc: 49.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050


Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM




#21 Endrin ketone

R.T.: 7.651 min
Delta R.T.: -0.006 min
Response: 166146342
Conc: 53.32 ng/ml




#22 Mirex

R.T.: 7.113 min
Delta R.T.: -0.007 min
Response: 477191773
Conc: 46.16 ng/ml



#22 Mirex

R.T.: 8.100 min
Delta R.T.: -0.007 min
Response: 124047988
Conc: 49.06 ng/ml




#28 Decachlorobiphenyl

R.T.: 7.984 min
Delta R.T.: -0.008 min
Response: 559686719
Conc: 44.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
7/11/2019 8:18:09 AM



#28 Decachlorobiphenyl

R.T.: 8.965 min
Delta R.T.: -0.008 min
Response: 151257710
Conc: 45.06 ng/ml