

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071020\
 Data File : PL059802.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2020 17:44
 Operator : DD\AJ
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL071020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 18:47:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 10 18:46:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.465	4.013	115.8E6	142.9E6	51.513	51.536
28)	SA Decachlor...	8.231	9.104	85855789	107.4E6	50.728	51.499
Target Compounds							
23)	Chlordane-1	4.337	5.010	30976527	50324474	505.602	518.230
24)	Chlordane-2	4.840	5.479	31768951	57847214	501.758	504.884
25)	Chlordane-3	5.410	6.128	87460002	201.4E6	517.161	520.416
26)	Chlordane-4	5.468	6.203	72393481	241.3E6	518.877	521.244
27)	Chlordane-5	6.305	7.010	28954623	42106099	510.299	542.149



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071020\
Data File : PL059802.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2020 17:44
Operator : DD\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL071020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 18:47:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071020.M
Quant Title : GC Extractables
QLast Update : Fri Jul 10 18:46:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.465 min
Delta R.T.: 0.000 min
Response: 115806515
Conc: 51.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL071020




#1 Tetrachloro-m-xylene

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 142911928
Conc: 51.54 ng/ml




#23 Chlordane-1

R.T.: 4.337 min
Delta R.T.: 0.000 min
Response: 30976527
Conc: 505.60 ng/ml



#23 Chlordane-1


R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 50324474
Conc: 518.23 ng/ml



#24 Chlordane-2


R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 31768951
Conc: 501.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL071020




#24 Chlordane-2

R.T.: 5.479 min
Delta R.T.: 0.000 min
Response: 57847214
Conc: 504.88 ng/ml




#25 Chlordane-3

R.T.: 5.410 min
Delta R.T.: 0.000 min
Response: 87460002
Conc: 517.16 ng/ml



#25 Chlordane-3


R.T.: 6.128 min
Delta R.T.: 0.000 min
Response: 201358311
Conc: 520.42 ng/ml



#26 Chlordane-4


R.T.: 5.468 min
Delta R.T.: 0.000 min
Response: 72393481
Conc: 518.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL071020




#26 Chlordane-4

R.T.: 6.203 min
Delta R.T.: 0.000 min
Response: 241283313
Conc: 521.24 ng/ml




#27 Chlordane-5

R.T.: 6.305 min
Delta R.T.: 0.000 min
Response: 28954623
Conc: 510.30 ng/ml



#27 Chlordane-5

R.T.: 7.010 min
Delta R.T.: 0.000 min
Response: 42106099
Conc: 542.15 ng/ml



#28 Decachlorobiphenyl


R.T.: 8.231 min

Delta R.T.: 0.000 min

Response: 85855789

Conc: 50.73 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL071020



#28 Decachlorobiphenyl

R.T.: 9.104 min

Delta R.T.: 0.000 min

Response: 107414389

Conc: 51.50 ng/ml