

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071119\
 Data File : PL050323.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2019 03:31
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/12/2019 10:10:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 04:30:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.322	3.932	626.1E6	175.9E6	55.518	57.464
28)	SA Decachlor...	7.987	8.966	540.7E6	157.7E6	42.552	46.984
Target Compounds							
2)	A alpha-BHC	3.749	4.321	1055.2E6	260.4E6	56.164	58.163
3)	MA gamma-BHC...	4.025	4.604	974.0E6	243.3E6	54.932	58.662
4)	MA Heptachlor	4.309	5.106	1010.7E6	252.4E6	55.604	58.316
5)	MB Aldrin	4.547	5.409	958.2E6	235.2E6	55.489	60.032
6)	B beta-BHC	4.279	4.780	403.9E6	109.0E6	55.216	56.271
7)	B delta-BHC	4.473	4.994	999.5E6	238.6E6	55.888	60.578
8)	B Heptachlor...	4.985	5.796	883.9E6	217.6E6	54.909	59.843
9)	A Endosulfan I	5.315	6.151	825.3E6	196.8E6	53.337	57.934
10)	B gamma-Chl...	5.207	6.030	927.5E6	217.7E6	54.274	58.431
11)	B alpha-Chl...	5.263	6.101	891.6E6	213.5E6	53.715	57.716
12)	B 4,4'-DDE	5.430	6.262	674.7E6	207.0E6	40.503	59.640 #
13)	MA Dieldrin	5.554	6.408	905.0E6	205.9E6	54.328	58.031
14)	MA Endrin	5.806	6.623	799.8E6	170.2E6	49.940	53.906m
15)	B Endosulfa...	6.082	6.837	718.0E6	172.8E6	49.275	53.638
16)	A 4,4'-DDD	5.941	6.751	700.6E6	156.4E6	55.494	59.832m
17)	MA 4,4'-DDT	6.175	7.050	603.1E6	140.1E6	45.430	47.849
18)	B Endrin al...	6.248	6.961	583.1E6	144.9E6	49.366	52.778
19)	B Endosulfa...	6.460	7.184	631.6E6	152.3E6	48.712	50.593
20)	A Methoxychlor	6.720	7.510	306.6E6	79437177	47.032	47.659
21)	B Endrin ke...	6.942	7.652	638.2E6	156.6E6	45.403	50.248
22)	Mirex	7.115	8.102	455.5E6	127.6E6	44.068	50.444

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071119\
 Data File : PL050323.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2019 03:31
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/12/2019 10:10:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 04:30:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

