

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071122\
 Data File : PL076360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2022 11:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 23:54:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 08 16:50:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.646	4.628	35544096	87741470	49.340	43.396
28)	SA Decachlor...	11.636	10.367	29066773	69477183	49.269	41.407
Target Compounds							
2)	A alpha-BHC	6.229	5.330	46426292	116.0E6	49.536	42.305
3)	MA gamma-BHC...	6.628	5.750	43586398	107.7E6	49.736	43.747
4)	MA Heptachlor	7.282	6.157	42547847	102.5E6	49.074	42.250
5)	MB Aldrin	7.655	6.481	39114103	95319555	50.349	43.970
6)	B beta-BHC	6.881	6.130	24531728	45731062	55.810	41.815 #
7)	B delta-BHC	7.150	6.390	37759974	103.6E6	56.392	43.307
8)	B Heptachlo...	8.123	7.055	35690233	86370584	45.844	43.821
9)	A Endosulfan I	8.531	7.456	33066643	79673625	48.953	43.206
10)	B gamma-Chl...	8.393	7.331	35336892	87881811	48.625	44.072
11)	B alpha-Chl...	8.474	7.400	35483625	86956867	49.031	43.250
12)	B 4,4'-DDE	8.660	7.612	30870484	79505929	49.797	43.257
13)	MA Dieldrin	8.822	7.741	33436149	81730849	49.493	43.205
14)	MA Endrin	9.062	8.031	27640766	71045066	48.039	49.142
15)	B Endosulfa...	9.296	8.342	28443024	66581993	47.608	43.081
16)	A 4,4'-DDD	9.202	8.196	25641759	62484787	49.078	43.094
17)	MA 4,4'-DDT	9.521	8.455	28749582	67713643	48.394	43.310
18)	B Endrin al...	9.431	8.529	23468846	54061355	48.625	43.142
19)	B Endosulfa...	9.669	8.759	28458157	68411125	47.668	43.020
20)	A Methoxychlor	10.006	9.047	15949275	36039843	47.594	42.248
21)	B Endrin ke...	10.170	9.277	32964070	75501600	48.760	44.866
22)	Mirex	10.643	9.462	24165681	58881098	47.251	44.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071122\
Data File : PL076360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2022 11:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 23:54:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
Quant Title : GC Extractables
QLast Update : Fri Jul 08 16:50:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

