

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071218\
 Data File : PL037920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2018 12:47
 Operator : AJ\SJ
 Sample : J3828-05 5X
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 EO-01-071018-C

Manual Integrations
 APPROVED

Sohil
 7/13/2018 2:33:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 13:26:59 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
 Quant Title : GC Extractables
 QLast Update : Thu Jun 21 02:27:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.332	3.934	15823946	4370500	2.000	1.580
28) SA Decachlor...	8.008	8.944	14393142	5525977	2.385	2.785m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071218\
Data File : PL037920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2018 12:47
Operator : AJ\SJ
Sample : J3828-05 5X
Misc :
ALS Vial : 42 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
EO-01-071018-C


Manual Integrations
APPROVED

Sohil
7/13/2018 2:33:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 13:26:59 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_L\methods\PL062018.M
Quant Title : GC Extractables
QLast Update : Thu Jun 21 02:27:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.332 min
Delta R.T.: -0.005 min
Response: 15823946
Conc: 2.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
EO-01-071018-C


Manual Integrations
APPROVED

Sohil
7/13/2018 2:33:01 PM




#1 Tetrachloro-m-xylene

R.T.: 3.934 min
Delta R.T.: -0.003 min
Response: 4370500
Conc: 1.58 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.008 min
Delta R.T.: -0.013 min
Response: 14393142
Conc: 2.39 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.944 min
Delta R.T.: -0.007 min
Response: 5525977
Conc: 2.79 ng/ml m