

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071219\
 Data File : PL050376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2019 02:56
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 7/16/2019 10:14:10 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 04:54:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 21 02:23:20 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.318	3.931	565.3E6	169.9E6	50.123	55.504
28)	SA Decachlor...	7.981	8.962	610.4E6	173.2E6	48.039	51.598
Target Compounds							
2)	A alpha-BHC	3.746	4.319	941.5E6	249.9E6	50.113	55.821
3)	MA gamma-BHC...	4.022	4.602	877.0E6	232.8E6	49.465	56.123
4)	MA Heptachlor	4.305	5.104	892.6E6	233.1E6	49.107	53.865
5)	MB Aldrin	4.544	5.406	904.4E6	226.2E6	52.377	57.748
6)	B beta-BHC	4.276	4.778	370.0E6	106.0E6	50.581	54.722
7)	B delta-BHC	4.468	4.992	903.7E6	228.0E6	50.534m	57.891
8)	B Heptachlor...	4.980	5.793	778.5E6	216.9E6	48.361	59.670
9)	A Endosulfan I	5.311	6.148	775.1E6	199.1E6	50.098	58.612
10)	B gamma-Chl...	5.202	6.027	868.8E6	218.4E6	50.841	58.612
11)	B alpha-Chl...	5.258	6.099	843.6E6	212.8E6	50.822	57.540
12)	B 4,4'-DDE	5.426	6.259	676.9E6	209.9E6	40.630	60.485 #
13)	MA Dieldrin	5.550	6.404	856.5E6	209.7E6	51.419	59.119
14)	MA Endrin	5.801	6.620	807.1E6	173.6E6	50.399m	54.974m
15)	B Endosulfa...	6.077	6.834	784.4E6	183.4E6	53.830	56.946
16)	A 4,4'-DDD	5.936	6.748	670.0E6	159.9E6	53.072m	61.167m
17)	MA 4,4'-DDT	6.170	7.046	678.7E6	155.9E6	51.126	53.234
18)	B Endrin al...	6.244	6.958	622.0E6	154.5E6	52.657	56.252
19)	B Endosulfa...	6.455	7.180	625.3E6	160.1E6	48.231	53.207
20)	A Methoxychlor	6.715	7.507	314.6E6	86751456	48.264	52.047
21)	B Endrin ke...	6.937	7.648	693.8E6	172.7E6	49.364	55.435
22)	Mirex	7.110	8.098	516.2E6	136.4E6	49.937	53.950

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071219\
Data File : PL050376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2019 02:56
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
7/16/2019 10:14:10 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 04:54:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL062019.M
Quant Title : GC Extractables
QLast Update : Fri Jun 21 02:23:20 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

