

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071222\
 Data File : PL076376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2022 10:25
 Operator : AR\AJ
 Sample : I.BLK (Sig #1); L.BLK (Sig #2)
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/13/2022
 Supervised By : Ankita Jodhani 07/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 14:11:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 08 16:50:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.642	4.627	15018707	38184916	20.848	18.886
2)	SA Decachlor...	11.633	10.367	9517172	29188301	16.132m	17.395

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071222\
Data File : PL076376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2022 10:25
Operator : AR\AJ
Sample : I.BLK (Sig #1); L.BLK (Sig #2)
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/13/2022
Supervised By : Ankita Jodhani 07/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 14:11:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
Quant Title : GC Extractables
QLast Update : Fri Jul 08 16:50:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 15018707
Conc: 20.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/13/2022
Supervised By :Ankita Jodhani 07/13/2022




#1 Tetrachloro-m-xylene

R.T.: 4.627 min
Delta R.T.: 0.000 min
Response: 38184916
Conc: 18.89 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.633 min
Delta R.T.: 0.001 min
Response: 9517172
Conc: 16.13 ng/ml m



#28 Decachlorobiphenyl

R.T.: 10.367 min
Delta R.T.: 0.000 min
Response: 29188301
Conc: 17.40 ng/ml