

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX071318.M
 Title : GC Extractables
 Last Update : Sat Jul 14 00:33:14 2018
 Response Via : Initial Calibration

Calibration Files

500 =PL037982.D 1000 =PL037980.D 250 =PL037983.D
 750 =PL037981.D 50 =PL037984.D

Compound			500	1000	250	750	50	Avg	%RSD	

1) SA	Tetrachloro-m-xylene		7.585	7.536	7.375	7.542	7.125	7.433	E6	2.55
2)	Toxaphene-1		4.885	5.030	4.973	4.979	5.088	4.991	E4	1.50
3)	Toxaphene-2		1.263	1.309	1.240	1.281	1.129	1.244	E5	5.57
4)	Toxaphene-3		9.211	9.841	8.837	9.386	7.666	8.988	E4	9.15
5)	Toxaphene-4		4.630	4.782	4.465	4.651	4.550	4.615	E4	2.56
6)	Toxaphene-5		1.523	1.575	1.500	1.522	1.161	1.456	E5	11.49
7) SA	Decachlorobiphenyl		7.630	7.376	7.809	7.406	8.127	7.670	E6	4.05

Signal #2 Calibration Files

500 =PL037982.D 1000 =PL037980.D 250 =PL037983.D
 750 =PL037981.D 50 =PL037984.D

Compound			500	1000	250	750	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		2.780	2.780	2.711	2.758	2.369	2.679	E6	6.56
2)	Toxaphene-1		1.301	1.262	1.297	1.260	1.360	1.296	E4	3.14
3)	Toxaphene-2		1.750	1.896	1.782	1.795	1.726	1.790	E4	3.64
4)	Toxaphene-3		6.052	6.152	5.863	6.056	4.945	5.814	E4	8.54
5)	Toxaphene-4		3.467	3.617	3.570	3.569	3.304	3.505	E4	3.58
6)	Toxaphene-5		1.447	1.455	1.400	1.392	1.425	1.424	E4	1.92
7) SA	Decachlorobiphenyl		2.475	2.415	2.533	2.431	2.553	2.481	E6	2.46

(#) = Out of Range