

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071321\
 Data File : PL068099.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2021 15:46
 Operator : AR\AJ
 Sample : PB137645BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 05:07:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071221.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 14 04:57:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.673	5.582	10556346	20347610	19.073	19.897
28)	SA Decachlor...	10.479	11.509	10988209	16550065	19.800	21.238
Target Compounds							
2)	A alpha-BHC	5.394	0.000	235674	0	0.323	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071321\
Data File : PL068099.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2021 15:46
Operator : AR\AJ
Sample : PB137645BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 05:07:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL071221.M
Quant Title : GC Extractables
QLast Update : Wed Jul 14 04:57:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.673 min
Delta R.T.: -0.003 min
Response: 10556346
Conc: 19.07 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.582 min
Delta R.T.: -0.002 min
Response: 20347610
Conc: 19.90 ng/ml




#2 alpha-BHC

R.T.: 5.394 min
Delta R.T.: 0.010 min
Response: 235674
Conc: 0.32 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T.: 6.165 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.479 min
Delta R.T.: -0.006 min
Response: 10988209
Conc: 19.80 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 11.509 min
Delta R.T.: -0.003 min
Response: 16550065
Conc: 21.24 ng/ml