

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071322\
 Data File : PL076423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 14:38
 Operator : AR\AJ
 Sample : L.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 00:22:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 08 16:50:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.640	4.625	13435266	37568681	18.650	18.581
28)	SA Decachlor...	11.631	10.363	12664467	33171538	21.467	19.769
Target Compounds							
2)	A alpha-BHC	0.000	5.334	0	703976	N.D.	0.257 #
7)	B delta-BHC	7.120f	6.392	363611	514842	0.543	0.215 #
11)	B alpha-Chl...	0.000	7.373f	0	198928	N.D.	0.099 #
13)	MA Dieldrin	0.000	7.734	0	250902	N.D.	0.133 #
14)	MA Endrin	0.000	8.050f	0	156718	N.D.	0.108 #
15)	B Endosulfa...	9.301	0.000	263303	0	0.441	N.D. #
17)	MA 4,4'-DDT	0.000	8.448	0	239327	N.D.	0.153 #
21)	B Endrin ke...	10.182f	0.000	531138	0	0.786	N.D. #
26)	Chlordane-4	0.000	7.373f	0	198928	N.D.	1.079 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071322\
Data File : PL076423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 14:38
Operator : AR\AJ
Sample : L.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 00:22:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
Quant Title : GC Extractables
QLast Update : Fri Jul 08 16:50:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 13435266
Conc: 18.65 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.625 min
Delta R.T.: -0.003 min
Response: 37568681
Conc: 18.58 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 6.225 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 5.334 min
Delta R.T.: 0.005 min
Response: 703976
Conc: 0.26 ng/ml



#7 delta-BHC


R.T.: 7.120 min
Delta R.T.: -0.026 min
Response: 363611
Conc: 0.54 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 6.392 min
Delta R.T.: 0.003 min
Response: 514842
Conc: 0.22 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 8.471 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 7.373 min
Delta R.T.: -0.025 min
Response: 198928
Conc: 0.10 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T. : 8.817 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 7.734 min
Delta R.T.: -0.005 min
Response: 250902
Conc: 0.13 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 9.058 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 8.050 min
Delta R.T.: 0.020 min
Response: 156718
Conc: 0.11 ng/ml



#15 Endosulfan II


R.T.: 9.301 min
Delta R.T.: 0.010 min
Response: 263303
Conc: 0.44 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.340 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 9.517 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 8.448 min
Delta R.T.: -0.005 min
Response: 239327
Conc: 0.15 ng/ml



#21 Endrin ketone


R.T.: 10.182 min
Delta R.T.: 0.015 min
Response: 531138
Conc: 0.79 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 9.275 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 8.472 min
Response: 0
Conc: N.D.



#26 Chlordane-4


R.T.: 7.373 min
Delta R.T.: -0.022 min
Response: 198928
Conc: 1.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.631 min
Delta R.T.: 0.000 min
Response: 12664467
Conc: 21.47 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 10.363 min
Delta R.T.: -0.004 min
Response: 33171538
Conc: 19.77 ng/ml