

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071422\
 Data File : PL076443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2022 15:04
 Operator : AR\AJ
 Sample : N3711-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSC124035

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/15/2022
 Supervised By :Ankita Jodhani 07/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 00:11:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 08 16:50:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.640	4.625	7167482	21843278	9.949	10.804
28) SA Decachlor...	11.631	10.362	3812121	11104279	6.462	6.618
Target Compounds						
17) MA 4,4'-DDT	9.515	8.448	314066	1331680	0.529m	0.852 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL071422\
Data File : PL076443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2022 15:04
Operator : AR\AJ
Sample : N3711-05
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PSC124035


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/15/2022
Supervised By : Ankita Jodhani 07/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 00:11:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL070822.M
Quant Title : GC Extractables
QLast Update : Fri Jul 08 16:50:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 7167482
Conc: 9.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSC124035


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/15/2022
Supervised By :Ankita Jodhani 07/15/2022




#1 Tetrachloro-m-xylene

R.T.: 4.625 min
Delta R.T.: -0.003 min
Response: 21843278
Conc: 10.80 ng/ml




#17 4,4'-DDT

R.T.: 9.515 min
Delta R.T.: -0.002 min
Response: 314066
Conc: 0.53 ng/ml m



#17 4,4'-DDT

R.T.: 8.448 min
Delta R.T.: -0.006 min
Response: 1331680
Conc: 0.85 ng/ml




#28 Decachlorobiphenyl

R.T.: 11.631 min
Delta R.T.: 0.000 min
Response: 3812121
Conc: 6.46 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSC124035

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/15/2022
Supervised By :Ankita Jodhani 07/15/2022



#28 Decachlorobiphenyl

R.T.: 10.362 min
Delta R.T.: -0.005 min
Response: 11104279
Conc: 6.62 ng/ml